

RESEARCH

Open Access



Peak detection in intracranial pressure signal waveforms: a comparative study

Miaomiao Wei¹, Solventa Krakauskaitė³, Sreya Subramanian³ and Fabien Scalzo^{2,3*}

*Correspondence:
fabien.scalzo@pepperdine.edu

¹ Department of Electronic and Information, Zhongyuan University of Technology, Zhengzhou, China

² Department of Neurology, University of California, Los Angeles (UCLA), Los Angeles, USA

³ Keck Data Science Institute, Pepperdine University, Malibu, USA

Abstract

Background: The monitoring and analysis of quasi-periodic biological signals such as electrocardiography (ECG), intracranial pressure (ICP), and cerebral blood flow velocity (CBFV) waveforms plays an important role in the early detection of adverse patient events and contributes to improved care management in the intensive care unit (ICU). This work quantitatively evaluates existing computational frameworks for automatically extracting peaks within ICP waveforms.

Methods: Peak detection techniques based on state-of-the-art machine learning models were evaluated in terms of robustness to varying noise levels. The evaluation was performed on a dataset of ICP signals assembled from 700 h of monitoring from 64 neurosurgical patients. The groundtruth of the peak locations was established manually on a subset of 13,611 pulses. Additional evaluation was performed using a simulated dataset of ICP with controlled temporal dynamics and noise.

Results: The quantitative analysis of peak detection algorithms applied to individual waveforms indicates that most techniques provide acceptable accuracy with a mean absolute error (MAE) ≤ 10 ms without noise. In the presence of a higher noise level, however, only kernel spectral regression and random forest remain below that error threshold while the performance of other techniques deteriorates. Our experiments also demonstrated that tracking methods such as Bayesian inference and long short-term memory (LSTM) can be applied continuously and provide additional robustness in situations where single pulse analysis methods fail, such as missing data.

Conclusion: While machine learning-based peak detection methods require manually labeled data for training, these models outperform conventional signal processing ones based on handcrafted rules and should be considered for peak detection in modern frameworks. In particular, peak tracking methods that incorporate temporal information between successive periods of the signals have demonstrated in our experiments to provide more robustness to noise and temporary artifacts that commonly arise as part of the monitoring setup in the clinical setting.

Introduction

The monitoring of quasi-periodic biological signals such as arterial blood pressure (ABP), intracranial pressure (ICP), and electrocardiography (ECG) plays a fundamental role in the study of numerous disorders and diseases. These biological signals have



© The Author(s) 2024. **Open Access** This article is licensed under a Creative Commons Attribution 4.0 International License, which permits use, sharing, adaptation, distribution and reproduction in any medium or format, as long as you give appropriate credit to the original author(s) and the source, provide a link to the Creative Commons licence, and indicate if changes were made. The images or other third party material in this article are included in the article's Creative Commons licence, unless indicated otherwise in a credit line to the material. If material is not included in the article's Creative Commons licence and your intended use is not permitted by statutory regulation or exceeds the permitted use, you will need to obtain permission directly from the copyright holder. To view a copy of this licence, visit <http://creativecommons.org/licenses/by/4.0/>. The Creative Commons Public Domain Dedication waiver (<http://creativecommons.org/publicdomain/zero/1.0/>) applies to the data made available in this article, unless otherwise stated in a credit line to the data.

something in common; they all exhibit characteristic variations that clinicians can use as markers of physiological change and provide additional insights through further analyses. This comparative study focuses on the ICP waveform, which is a quasi-periodic signal. As visualized in Fig. 1, each pulse can be associated with three peaks due to its triphasic nature [1]. Therefore, ICP morphological analysis often relies on identifying these three peaks. Based on their latency (i.e., time) and elevation (i.e., height), it is possible to characterize the morphology over time and then compute statistics of the ICP waveform for a particular time interval. We provide in this paper a comparative analysis of techniques for detecting the three peaks across the periods of the signal.

The study of the variations of the ICP signal is particularly important for patients of the NICU treated for traumatic injury (TBI) as the morphological variations observed in the ICP waveform may reveal the compensatory ability of the brain in the presence of cerebrovascular disruptions. Several computational frameworks have been developed to study how the changes observed in the morphology of the ICP signal are associated with the development of cerebral vasospasm [2], intracranial hypertension [3, 4], and abrupt changes in the cerebral blood carbon dioxide (CO₂) levels [1, 5], and changes in the craniospinal compliance [6]. In addition to the average change of the ICP, studies [7, 8] have linked the morphology of the ICP waveforms with the prognosis of patients with a head injury. Hence, exploring ICP morphological characteristics such as peaks may help monitor pathophysiological intracranial changes.

Traditionally, peak detection in biological signals has been achieved using signal processing methods, including threshold-based and filtering methods [9, 10]. More recently, machine learning (ML)-based methods [11, 12] have been developed to solve this problem. ML methods are usually built on top of signal processing methods to obtain more robustness to noise by capturing the characteristics of the peaks and adapting to the noise profile of the signal specific to the context in which it is acquired. Many machine learning models are available, including neural networks, random forests, support vector machines (SVM), long short-term memory (LSTM) [13], etc. In the context of ICP analysis, these models can be grouped into two main categories depending on whether they are processing a single pulse at a time or if they are processing the continuous signal instead. It remains unclear, however, which methods are the best-performing ones on ICP signals.




Fig. 1 Example of continuous real-time monitoring of ICP waveform from a patient included in our dataset. Red dots depict the three peaks within each triphasic waveform

This study compares ICP peak detection methods. After a technical review and description of the literature, we describe our peak detection experiments performed on actual and simulated ICP datasets.

State-of-the-art

Summary

Peak detection techniques on quasi-periodic signals can be grouped into two categories depending on whether they are processing individual waveforms or utilizing temporal patterns from previous waveforms to identify peaks in the current waveform. Traditionally, peak detection on individual waveforms has been achieved using signal processing techniques. In that scenario, a single beat/period of the signal waveform is used as input to a model, and the output corresponds to the latency and/or elevation of the peaks within that beat. It is common to assume that a peak is simply a local extrema of the curvature of the signal. In the context of peak detection in biomedical signals such as ICP, however, artifacts and noise make peak recognition challenging when relying on pre-defined heuristics only and often result in false positive detections. To tackle this problem, peak detection techniques have been improved using data-driven approaches (i.e., machine learning) that utilize a training set of data samples to learn a peak detection model. Data-driven techniques have demonstrated significant promise to improve the robustness of peak detection in biomedical signals.

Because strong correlations may exist between the peak locations of successive waveforms, tracking techniques such as Kalman filter, long short-term memory (LSTM), Bayesian inference, or MOCAIP (Morphological Clustering and Analysis of ICP Pulse) [14] can capture the temporal information between successive pulses to refine the location of the peaks. We review in the following sub-sections the techniques available in the literature to perform peak detection on individual (Section II-B) and continuous (Section II-C) waveforms.

Peak detection on individual waveforms

Peak detection is assumed here to be performed on individual ICP pulses previously segmented from continuous ICP waveforms [10, 15]. In particular, we divide the single waveform methods into two sub-groups depending on whether they are based on signal processing only or if they also utilize data-driven models during processing.

Signal-processing techniques

In signal processing, it is common to think of peak detection as a search for a local extrema in the curvature of the signal. Most of signal-based methods utilize the local structure of the signal to identify the peaks. Among them, we identify threshold-based processes [16, 17], derivative-based techniques [18], and transform domain techniques [19–21]. Other methods perform peak detection by incorporating a larger context to describe the signature of each peak within the beat, which includes intensity weighted variance [22], filter-based techniques [23, 24], histogram-based techniques [25, 26], techniques using entropy [27], momentum [28], stochastic resonance [29], higher-order statistics [30], nonlinear energy operator [31], empirical mode decomposition [32]. More

advanced techniques such as the wavelet transform and entropy of coefficients [33, 34] have also achieved promising results.

While peak detection techniques based on signal processing perform well on a wide range of applications, they encounter significant challenges when applied to real-world ICP data due to the variability across subjects, motion artifacts, and hardware acquisition noise characteristic of ICP waveforms. As described in the next subsection, these challenges have pushed researchers to utilize more robust techniques for these variations.

Data-driven techniques

Data-driven techniques utilize training data samples to infer a model optimized for peak detection. While the learning algorithm algorithms can take many different forms, many approaches formalize peak detection as a regression analysis problem between the input ICP waveform and the location of each peak (as the target output). This section gives an overview of ML methods used to detect ICP peaks and will be evaluated in our experiments. These methods include spectral regression (SR) [35], neural networks (NN) [36], support vector machines (SVM) [37], and extremely randomized decision trees (Extra-Trees) [38].

Spectral regression (SR). The SR algorithm [35] is a nonlinear regression method incorporating graph-based analysis with regularized linear regression. Assuming a set of N input data samples $\{x_0, x_1, \dots, x_{N-1}\}$ and their corresponding predicted output $\{\hat{y}_0, \hat{y}_1, \dots, \hat{y}_{N-1}\}$, the objective is to learn a regression model that outputs similar predictions $\hat{y}_i$ for input samples x_i near each other in a graph representation. The regression model is obtained by minimizing the following measure ϕ :

$$\phi = \sum_{i,j=1}^N (\hat{y}_i - \hat{y}_j)^2 W_{i,j}, \quad (1)$$

where $W_{i,j}$ is the affinity matrix $W \in \mathbb{R}^{N \times N}$ that assigns a value to $W_{i,j}$ to indicate the similarity of the two input samples x_i, x_j ; where i, j are used to represent the index of the i^{th} and j^{th} data samples, respectively.

While SR has been developed to solve linear problems, it can be extended to nonlinear problems using a kernel projection, which projects the original observation x_i into a higher dimension using a nonlinear kernel. In the kernel version of SR, referred to as kernel spectral regression (KSR), the data input samples x_i are replaced by the projected vectors in Eq.(2). In this study, we use the radial basis function (RBF) kernel:

$$R(x_i, x) = \exp(-\beta \|x_i - x\|^2), \quad \beta > 0. \quad (2)$$

Neural networks (NN). Neural networks is another popular machine learning model that can infer peak locations from an ICP waveform. Numerous neural network architectures exist; we focus here on a feed-forward network that comprises input, hidden, and output layers. The Levenberg–Marquardt algorithm [39] was used for its efficiency in training moderate-sized NNs. The SSE is used as the fitness function.

Support vector machine (SVM). A support vector machine (SVM) [37] is a supervised learning method that constructs a set of hyperplanes in a high-dimensional

space. SVM has been proven to be an effective tool in real-value function estimation. In the context of regression, SVM (also called support vector regression (SVR)) uses a n -dimensional tube to fit the data. During learning, the optimization process adopts an ϵ -insensitive loss function, penalizing predictions farther than the threshold ϵ from the desired output. The value of ϵ determines the diameter of the tube; a smaller value indicates a lower tolerance for error and affects the smoothness of the overall predictions. For regression problems, SVM aims to identify the parameters of a set of hyperplane(s)/tube(s) that best fit the data using the following metric:

$$\sum_{s \in SV}^n (\alpha_s^+ - \alpha_s^-) R(x_s, x) + b, \quad (3)$$

where SV is a subset of the input data samples x called “support vectors”, α^+ , α^- represents the learned dual coefficients, $R(x_s, x)$ is the response of the RBF kernel (Eq. 2) of the data sample x_s , and b is the bias.

Extremely randomized decision trees (extra-trees). Extra-Trees [38] is a regression method based on an ensemble of randomized decision trees. The learning of a randomized decision tree is performed by starting at the tree’s root node and successively splitting its left and right sub-trees. Each split (i.e., threshold) is obtained by sampling according to a Gaussian distribution estimated from the training samples. The process is repeated until a node has constant output values for all the training inputs. By building many randomized decision trees (e.g., $N > 100$), the model can make predictions by using a new input through each tree and computing the average prediction across all the trees.

Peak tracking on continuous waveforms

Although peak detection on individual pulses can achieve reasonable accuracy by identifying the signal signature of these peaks or learning a regression model between the waveform and the peak location with machine learning, processing pulses individually has some limitations. Hardware noise and human disturbance (such as motion artifacts) are inevitable in a clinical environment. These may cause distortion or even temporary loss of ICP waveform, making detecting and tracking peaks based solely on a single pulse challenging. Achieving continuous and real-time analysis of ICP waveforms is a high-level requirement of ICP monitoring in the NICU. Here, we describe techniques developed to process the continuous ICP signal and locate the peaks using temporal properties between successive beats as prior information, effectively tracking them across different periods. In the following, we describe Kalman filtering [40], Bayesian tracking [41], and LSTM (long short-term memory) [13], and MOCAIP (Morphological Clustering and Analysis of ICP Pulse) [14].

Kalman filtering. The Kalman filter algorithm [42] is a recursive algorithm that estimates the distribution of unknown variables from the measured noisy data. After several iterations, the estimated value is expected to converge to the actual value of the unknown variables; the location of the peaks in our case. The process is efficient as it only needs the current measured input, the previous state, and the uncertainty

state matrix to calculate the predicted value when the subsequent measurement is observed. The Kalman filter is composed of a prediction and an updating step.

The state variable $\hat{x}_k^-$ and its covariance P_k are estimated during the prediction step:

$$\hat{x}_k^- = A\hat{x}_{k-1} + Bu_k, \quad (4)$$

$$P_k^- = AP_{k-1}A^T + Q, \quad (5)$$

where A is the state-transition matrix, B is the control-input model, and Q represents the covariance of the noise.

During the updating step, these estimates are evaluated using a weighted average, such that a greater weight is set to estimations with greater confidence:

$$K_k = P_k^- H^T (HP_k^- H^T + R)^{-1}, \quad (6)$$

$$\hat{x}_k = \hat{x}_k^- + K_k(z_k - H\hat{x}_k^-), \quad (7)$$

$$P_k = (I - K_k H)P_k^-, \quad (8)$$

where H represents the observation matrix, $\hat{x}_k^-$ and $\hat{x}_k$ are the prior and posterior state estimates at step k . R is the measurement error covariance, z_k and u_k are the measurements and the control vectors at step k , and K represents the Kalman filtering gain.

Bayesian tracking. Nonparametric belief propagation (NBP) [43] is a probabilistic inference algorithm applied in computer vision to track the movements of people, animals, robots, cars, etc. We previously used NBP [41] to track ICP peaks in real-time. Bayesian inference associates continuous probability distributions as the location of each peak.

During detection, NBP utilizes a dynamic graph where nodes represent the location of each peak. Information between the different peaks of a current pulse, and between the peaks at the prior time point are propagated in the graph via a message-passing algorithm called Belief propagation. At the n th iteration, the message m passed from node a to b is expressed as:

$$m_{a,b}^n(h_a) \leftarrow \int \phi_{a,b}(h_a, h_b) \phi_a(h_a, o_a) \prod_{c \in C_{a \setminus b}} m_{c,a}^{n-1}(h_a) dh_a, \quad (9)$$

where $h_a \in \mathbf{h}$ represents the hidden variable at node a . $C_{a \setminus b}$ represents the set of nodes connected to a (except node b). $\phi_a(h_a, o_a)$ is the observation potential between hidden variable h_a and observation variable o_a of node a , $\phi_{a,b}(h_a, h_b)$ is the compatibility potential between hidden variables h_a and h_b . After several iterations, the approximation of n^{th} iteration $\hat{p}^n(h_a|o)$ converges to the true marginal distribution $p(h_a|o)$ is:

$$p(h_a|o) \sim \hat{p}^n(h_a|o) \leftarrow \phi_a(h_a, o_a) \prod_{b \in C_a} m_{a,b}^n(h_a). \quad (10)$$

In NBP, the message $m_{a,b}(h_a)$ is expressed as a mixture of D kernels:

$$m_{a,b}(h_a) = \frac{1}{D} \sum_{i=1}^D \omega_a^i N(h_a; \mu_a^i, \Sigma_a^i), \quad (11)$$

where ω_a^i is the weight of the i th kernel with mean μ_a^i and variance Σ_a^i . D is the number of particles used for estimation. The observation potential is represented as weighted mixtures of Gaussian density functions.

Long short time memory (LSTM). LSTM [13] is a type of recurrent neural network (RNN) that allows information to persist inside the network via loops in its architecture. LSTMs are particularly well suited to represent time series such as ICP waveforms. An LSTM cell is defined by a state that changes according to three types of gates:

- Input gates $\mathcal{I}_t \in \mathcal{R}^N$ update the state of the cell and decide which values should be updated.
- Forget gates $\mathcal{F}_t \in \mathcal{R}^N$ are used to select relevant information with respect to a previous state.
- Output gates $\mathcal{O}_t \in \mathcal{R}^N$ determine the final cell state and the output value.

Given an input sequence $x = \{x_1, x_2, \dots, x_T\}$ of length T with corresponding memory cell unit $C_t \in \mathcal{R}^N$ and hidden unit $h_t \in \mathcal{R}^N$ at time t , the parameters of the model are updated sequentially, as follows:

$$\mathcal{F}_t = \sigma(W_f.[h_{t-1}, x_t] + b_f), \quad (12)$$

$$\mathcal{I}_t = \sigma(W_i.[h_{t-1}, x_t] + b_i), \quad (13)$$

$$\mathcal{O}_t = \sigma(W_o[h_{t-1}, x_t] + b_o), \quad (14)$$

$$\tilde{C}_t = \tanh(W_c.[h_{t-1}, x_t] + b_c), \quad (15)$$

$$C_t = \mathcal{F}_t * C_{t-1} + \mathcal{I}_t * \tilde{C}_t, \quad (16)$$

$$h_t = \mathcal{O}_t * \tanh C_t. \quad (17)$$

The function $\sigma(x) = 1/(1 + e^{-x})$ used to compute $\mathcal{F}_t, \mathcal{I}_t, \mathcal{O}_t$ is a sigmoid function whose values lie within the range $[0, 1]$. In addition to input, forget, and output gates previously described, the LSTM makes use of a memory cell unit C_t obtained from the sum of the previous memory cell unit C_{t-1} modulated by $\mathcal{F}_t$, and a function of the current input x_t and previous hidden state h_{t-1} modulated by the input gate $\mathcal{I}_t$. The output gate $\mathcal{O}_t$ is then used to determine what parts should be considered and then multiplied with the tanh of the memory cell state C_t to produce the hidden unit h_t . By learning how much of the memory cell state C_t should be transferred to the hidden state h_t based on the input x_t and previous state, this structure allows the LSTM to capture complex temporal dynamics such as the ones present across ICP waveforms.

MOCAIP algorithm. The Morphological Clustering and Analysis of ICP Pulse (MOCAIP) [14] framework was designed to extract morphological variations of ICP

pulses. MOCAIP utilizes a Gaussian distribution as prior model for the peak location. The detection of ICP peaks is performed through the three main following steps:

- Pulse segmentation: The continuous ICP signal is segmented into a series of individual pulses using a dedicated algorithm [44] that utilizes ECG QRS markers [45]. A hierarchical clustering algorithm is utilized to extract a representative pulse over a segment of 1 min.
- Peak candidates detection. Candidate peaks are detected on the ICP pulse using its second derivative. They are extracted from the convex region and the concave region on the ascending edge of the signal or the concave part and the convex part on the falling edge of the signal.
- Peak Designation. The three peaks are selected from the set of candidate peaks such that they maximize the likelihood of belonging to a previously trained Gaussian mixture model (i.e., prior model).

Methods

Problem formulation

When acquired at a high enough frequency, ICP signals typically exhibit a sequence of waveform pulses such that each pulse includes three peaks, as illustrated in Fig. 2. We decompose the peak detection process on a raw signal by assuming that the continuous ICP waveform has been segmented into a set of individual beats $(s_1, s_2, \dots, s_n)$ using a standard beat segmentation algorithm [10, 15]. This is generally achieved with high accuracy - especially when the ECG signal is available. Assuming a segmented ICP waveform, we focus on two formulations of the peak detection problem. In the first case, we consider the task of detecting the three peaks within a single ICP pulse. In the second case, the peak detection is achieved by a tracking algorithm that exploits the estimated position of the peaks from previous pulses. In both formulations, a peak location is defined in terms of its temporal location $l \in \mathcal{R}$ and intracranial pressure elevation $e \in \mathcal{R}$, such that $p_{i \in 1,2,3} = \{l, e\}$ denotes the




Fig. 2 Illustration of a segmented pulse from a continuous ICP waveform and its three peaks (p1, p2, p3)

i th peak of the pulse. The goal is to obtain automatically the position of the peaks in each beat s_i using a peak detection algorithm P_d , which can be denoted as $P_d(s_i) = \{p_1, p_2, p_3\}$.

ICP data

Clinical dataset

The dataset of ICP signals used in this study was collected from 64 patients receiving treatment for various ICP-related disorders in the Neuro ICU. The ICP was acquired using intraparenchymal microsensors placed in the right frontal lobe. The raw ICP waveform was recorded continuously at a sample rate of either 240Hz or 400Hz. 153 segments of ICP signal lasting almost 5 h were extracted. ICP and ECG signal were then pre-processed to segment individual beats to produce a set of 14,230 raw pulses. Among them, 13,611 valid pulses were obtained and formed the clinical dataset used for our simulation. The dataset is particularly challenging because there is a large variability in the ICP signals due to each patient's condition.

In our experiments, the raw ICP waveforms were pre-processed before being used as input for single waveform or tracking analysis. The learning models used require a fixed length for the input data. ICP waveforms were first resampled to a fixed length because the waveforms' lengths are dependent on the patient's heart rate, which is variable. Each beat sample $\tilde{S}_i \in S$ was resampled to a vector of 400 values. A left shift was then performed to align the beats. We define the alignment point to be the minimum of each beat waveform:

$$\text{start} = \arg \min_i (\vec{Z}_i), \quad (18)$$

and perform a circular shift to set this point as the first element of each beat vector, where n is the length of $\vec{Z}_i$.

$$\vec{N}_i = \vec{Z}_{(i+\text{start})\bmod(n)}. \quad (19)$$

Since there is usually noise in ICP waveforms, which results in the distortion of the waveform, especially for sharp noise, the absolute magnitude of the waveform can be unreliable in clinical settings. To reduce this impact, each sample $\vec{X}_i$ is normalized so that its AUC is 1.

$$\vec{X}_{i,j} = \frac{\vec{X}_{i,j}}{\sum_{j=1}^J \vec{X}_{i,j}} \quad (20)$$

Three experienced researchers established the groundtruth by reviewing each ICP pulse and manually assigning the position of the three peaks. Specifically, the researcher's task was to select the suitable peak candidates for each peak (p_1 , p_2 , and p_3) among those automatically detected at curve inflections. Researchers cross-validated their results and, if necessary, harmonized them using the annotation of the previous and following pulses as reference. For a few difficult cases where the researchers could not agree on the position of some peaks, the pulse was removed from the dataset. This procedure ensured that the groundtruth is not biased to a specific researcher. A custom-made annotation tool allowed for flagging missing peaks. In our dataset, p_1 was missed in 1717 pulses, p_2 in 265 pulses, and p_3 in 34 pulses. Data from two patients were removed due to the

device malfunction. The data were acquired at the Ronald Reagan Medical Center at the University of California, Los Angeles (UCLA), and the UCLA Internal Review Board (IRB) approved the usage of this archived dataset.

Simulated dataset

To verify the effectiveness of the peak detection algorithms under controlled variability, we created a simulated dataset of ICP waveforms. A probabilistic generative model was used to simulate realistic shape variations of an ICP pulse. The model was formalized as a Gaussian Mixture Model (GMM) composed of three Gaussian components. The Gaussian Mixture model is a linear combination of Gaussian distributions:

$$p(x) = \sum_{k=1}^K \pi_k \mathcal{N}(x|\mu_k, \Sigma_k), \quad (21)$$

where π_k is the weight associated with the k^{th} component, and the number of components K was set to 3 in our experiments. The parameters π_k, μ_k, Σ_k were fitted using the clinical data using a random sample of 1, 500 ICP waveforms.

A series of pulses was then generated from this GMM model by incorporating an independent temporal change $c_{i \in 1,2,3} = \sin(z)$ on the mean of each component $\mu_k \in \mathcal{R}^2$. The model of the temporal dynamic was formalized as a sine wave function whose value c_k was added to its corresponding mean μ_k . It should be noted that two independent sine functions were used: one that acts on the latency and the other on the pressure of each peak. The generative model was then used to reconstruct individual waveform pulses at a sampling rate of 400Hz. The range of the sine wave was constrained by the fluctuation range observed in our clinical datasets. Figure 3 illustrates the variations induced by the sine wave on the latency of the three peaks.

Experiments

Our experiments aim to compare the accuracy of several machine-learning models in locating the peaks within the ICP signal. For both the clinical and simulated datasets, we evaluate the accuracy of the models in detecting the peaks under a varying amount of noise. We also perform evaluations to evaluate the robustness to missing data.




Fig. 3 Illustration of simulated ICP waveforms from our GMM generative model (a) and corresponding latency for the 3 peaks (p1, p2, p3) over a sequence of generated data (b)

Experiment #1: peak detection on clinical dataset

The evaluation performed as part of this experiment is carried out on individual waveforms where the input provided to the regression model does not include any context or waveforms from previous time points. The algorithms evaluated in our benchmark are spectral regression (SR), kernel spectral regression (KSR), neural networks (NN), support vector machines (SVM), and long short-term memory models (LSTM).

A tenfold cross-validation, performed at the patient level, is performed separately on the clinical and simulated datasets. For each training iteration, a threefold cross-validation is used on the training set to optimize the hyperparameters - this procedure is commonly referred to as a nested cross-validation. The ICP waveforms with missing peaks were included as part of the experiments. However, the missing peaks were ignored from the computation of the error. The input provided to the machine learning algorithms for both datasets is the ICP waveform re-sampled at 400Hz. The mean absolute error (MAE) and the root mean square error (RMSE) are used as a metric of accuracy and computed per peak and for each algorithm:

$$\text{MAE} = \frac{|y_i - \hat{y}_i|}{n}, \quad (22)$$

$$\text{RMSE} = \sqrt{\frac{\sum_{i=1}^n (y_i - \hat{y}_i)^2}{n}}, \quad (23)$$

where y_i represents the i th observation, $\hat{y}_i$ is the prediction of y_i for the given model, and n denotes the total number of observations. The average error is computed across the 3 peaks between the actual value of the peaks $y_i = (p_1, p_2, p_3)$ and the prediction $\hat{y}_i = (\hat{p}_1, \hat{p}_2, \hat{p}_3)$ of the regression method.

The monitoring of ICP can be adversely impacted by various noise and artifacts (including electromagnetic interference from other equipment and self-noise). In practice, it is manifested by abnormal fluctuations in the ICP waveform. To reflect these signal perturbations and evaluate the robustness of peak detection algorithm to them, we create noisy replication of our ICP datasets by adding varying uniform random noise levels (from 5 to 15% of the signal range) on the original ICP waveform.

Hyperparameters were optimized using nested cross-validation using only the training folds at each iteration. Specifically, we list below the optimized parameters for each method and list the implementation source. Matlab implementation of spectral regression and kernel spectral regression was obtained from Prof. Deng Cai's academic website at <http://www.cad.zju.edu.cn/home/dengcai/>. The spectral regression hyperparameters were the kernel type used for the affinity matrix W , the regularizer parameter α , and the number of neighbors used to compute W . For kernel spectral regression, an additional hyperparameter was used to control the standard deviation of the RBF kernel, which was also the optimized hyperparameter for SVM. For the neural network, the number of hidden layers/nodes, learning rate, optimizer were optimized. For the random forest, we optimized the number of decision trees. For LSTM, the number of hidden nodes was the only parameters fine-tuned. Except for spectral regression and kernel spectral regression (obtained from Deng Cai), the

implementation of all the methods obtained from Matlab official toolboxes as of version R2022a.

Experiment #2: peak detection simulated dataset

In this second experiment, the evaluation is conducted on a series of ICP pulses. In particular, we assume that a regression algorithm first predicts the position of the 3 peaks. Such prediction, affected by noise and artifacts, is then filtered using a tracking algorithm to obtain a refined position of the peaks by utilizing temporal dependencies between successive ICP pulses. The tracking algorithms evaluated are MOCAIP [14], nonparametric Bayesian tracking [41], Kalman filter [42], and LSTM [13]. The regression model used to obtain candidate peaks on single waveforms is KSR. Similarly to our previous experiment, we repeat the evaluation by adding various noise levels (5–15%) to the simulated data. The noise was uniformly distributed relative to the range of the data and added independently to the latency and elevation values.

All the tracking code was implemented in Matlab and executed under the version R2022a. MOCAIP and the Bayesian version of MOCAIP are available on GitHub under <https://github.com/NeuroResearchCore/trackLight>.

In some cases, the patient's movement or other physiological activities will cause a loss of connectivity, resulting in data loss in the ICP waveform. Without a signal, traditional ICP peak detection algorithms based on a single waveform will fail. We modified the simulated dataset to set some intervals to null to simulate this situation. This helps verify whether the tracking algorithm can utilize prior information to keep track of the peak over time. To simulate the missing ICP waveform, we divide the simulated waveform into several groups, and two or three missing segments of various lengths (2–4 pulses missing) are produced in each group to ensure the randomness of the missing situation and the dispersion of its distribution in the whole waveform.

Results

Experiment #1: peak detection on clinical dataset

The mean absolute error (MAE) of six peak detection algorithms on individual waveforms is reported in Table 1 after a tenfold cross-validation. The table summarizes the results for each of the three peaks. Each sub-table corresponds to the performance concerning one of the peaks. The fourth sub-table represents the average performance across all peaks. The columns correspond to the noise levels (0–15%). The MAE values (in milliseconds) were mapped to a color such that blue indicates lower error, and yellow indicates higher error.

On average, the results on the clinical dataset show that the error is the smallest for p_2 , followed by p_1 , and finally, p_3 . This is because the position of p_2 is more stable than other peaks. Without added noise, KSR, SVM, and Random forests perform best (RMSE = 0.08, MAE ≤ 4 ms) when considering the average of the three peaks. In the presence of 15% noise, the estimated error of KSR is the smallest (RMSE = 0.13, MAE ≤ 10 ms) as it appears to be less affected by noise. As expected, the MAE and RMSE of all algorithms increases due to the noise level. We note that not all algorithms grow at the same rate than the noise. For example, the error of spectral regression and neural networks increases much higher than in other methods.

Table 1 Performance of peak detection algorithms in terms of mean absolute error in milliseconds (ms) after evaluation on clinical ICP data with varying noise levels (from 0 to 15%)

Peak 1	Noise 0%	Noise 5%	Noise 10%	Noise 15%
Spectral Regression	5.6	6.4	8.8	11.8
Kernel Spectral Regression	3.9	4.4	5.6	7.2
SVM	3.5	4.9	9.0	11.0
LSTM	6.7	5.6	6.6	8.9
Neural Network	4.6	15.4	20.9	25.3
Random Forests	3.7	4.7	5.7	6.7
MOCAIP	9.6	12.6	15.0	16.6

Peak 2	Noise 0%	Noise 5%	Noise 10%	Noise 15%
Spectral Regression	6.7	18.2	34.9	51.8
Kernel Spectral Regression	3.6	4.3	5.7	7.5
SVM	3.6	5.7	10.5	12.9
LSTM	7.9	8.6	10.1	11.8
Neural Network	4.6	11.8	14.4	16.2
Random Forests	4.0	6.1	7.9	9.3
MOCAIP	19.3	37.8	42.4	43.7

Peak 3	Noise 0%	Noise 5%	Noise 10%	Noise 15%
Spectral Regression	10.4	17.2	28.7	41.1
Kernel Spectral Regression	7.0	8.1	10.3	13.2
SVM	6.3	10.1	20.4	25.5
LSTM	12.8	14.9	18.1	21.7
Neural Network	8.6	41.7	62.9	73.6
Random Forests	6.6	11.9	17.7	22.9
MOCAIP	31.2	54.4	64.9	62.4

All Peaks	Noise 0%	Noise 5%	Noise 10%	Noise 15%
Spectral Regression	7.5	13.9	24.1	34.9
Kernel Spectral Regression	4.8	5.6	7.2	9.3
SVM	4.5	6.9	13.3	16.5
LSTM	9.2	9.7	11.6	14.1
Neural Network	5.9	23.0	32.7	38.3
Random Forests	4.7	7.6	10.5	12.9
MOCAIP	20.0	35.0	40.7	40.9

Similarly, Table 2 provides the results after evaluating the peak detection methods on the simulated dataset. When the error is averaged over the three peaks, KSR offers the best performance among the six algorithms regardless of the noise level. The RMSE of spectral regression, LSTM, neural network, and random forests are higher ($RMSE \geq 0.29$, $MAE \leq 10$ ms) and are greatly affected by noise. From the results summarized in Tables 1 and 2, we conclude that KSR performs better than the algorithms when considering a single waveform at a time.

Experiment #2: peak detection on simulated dataset

The results of the peak tracking methods (Bayesian tracking, Kalman filter, LSTM, and MOCAIP) on continuous ICP are summarized in Table 2 and illustrated in Figs. 4 and 5. In Fig. 4, each plot includes the latency of the peak with noise (gray) and the filtered position of the tracking algorithm (color curves). These curves are repeated for each of the three peaks (p_1 , p_2 , and p_3), which can be judged according to the value range of its Y-axis. Figure 5 displays the results regarding the elevation of the first peak. For better visibility, we opted only to show the tracking of the first peak, as all peaks tend to be within the same elevation range in our simulated dataset.

Table 2 Performance of peak detection and tracking algorithms in terms of mean absolute error in milliseconds (ms) after evaluation on simulated ICP waveforms with varying noise levels (from 0 to 15%)

Peak 1		Noise 0%	Noise 5%	Noise 10%	Noise 15%
Single Waveform	Spectral Regression	12.7	17.3	23.4	30.4
	Kernel Spectral Regression	12.6	12.9	13.8	15.8
	SVM	12.8	11.5	13.1	16.4
	LSTM	13.4	13.4	13.4	13.5
	Neural Network	10.9	16.3	16.7	16.7
	Random Forests	13.7	14.0	14.2	14.2
	MOCAIP	4.6	41.2	36.6	34.4
Tracking	LSTM	1.1	1.1	1.7	1.9
	Bayesian	1.7	1.7	1.9	2.4
	Kalman Filter	0.0	1.4	2.4	3.6
	MOCAIP	2.0	2.2	2.8	3.6

Peak 2		Noise 0%	Noise 5%	Noise 10%	Noise 15%
Single Waveform	Spectral Regression	4.8	5.3	5.8	6.5
	Kernel Spectral Regression	3.6	3.5	3.5	4.1
	SVM	4.9	5.3	5.7	6.4
	LSTM	12.1	14.4	18.0	22.6
	Neural Network	13.4	16.4	16.3	16.6
	Random Forests	15.2	14.9	14.7	14.8
	MOCAIP	23.3	35.5	33.7	33.6
Tracking	LSTM	1.3	1.6	1.6	2.6
	Bayesian	1.8	1.8	1.9	2.0
	Kalman Filter	0.0	0.8	1.6	2.6
	MOCAIP	1.8	1.8	2.1	2.3

Peak 3		Noise 0%	Noise 5%	Noise 10%	Noise 15%
Single Waveform	Spectral Regression	6.9	6.8	6.8	6.9
	Kernel Spectral Regression	6.2	6.3	6.9	7.9
	SVM	7.2	7.3	7.6	8.1
	LSTM	8.9	8.9	8.9	9.0
	Neural Network	11.9	15.7	15.5	16.2
	Random Forests	13.2	13.3	13.7	14.1
	MOCAIP	26.8	29.9	29.5	30.0
Tracking	LSTM	1.4	1.7	1.6	1.7
	Bayesian	1.0	1.1	1.2	1.5
	Kalman Filter	0.0	0.5	1.0	1.5
	MOCAIP	0.9	1.0	1.1	1.3

All Peaks		Noise 0%	Noise 5%	Noise 10%	Noise 15%
Single Waveform	Spectral Regression	8.1	9.8	12.0	14.6
	Kernel Spectral Regression	7.5	7.6	8.1	9.2
	SVM	8.3	8.0	8.8	10.3
	LSTM	11.5	12.2	13.5	15.0
	Neural Network	12.1	16.2	16.2	16.5
	Random Forests	14.0	14.1	14.2	14.4
	MOCAIP	18.2	35.5	33.3	32.7
Tracking	LSTM	1.3	1.5	1.7	2.1
	Bayesian	1.5	1.6	1.7	2.0
	Kalman Filter	0.0	0.9	1.6	2.6
	MOCAIP	1.6	1.7	2.0	2.4

The results illustrated in Table 2 indicate that all tracking methods outperform single waveform techniques, especially in high noise. All tracking techniques perform equally well with a RMSE of 0.04–0.05 and $MAE \leq 3$ ms. This result is confirmed across the three peaks.

The accuracy of each tracking algorithm can be observed based on how close the estimate is to the groundtruth (shown in Fig. 3b). The tracking results of the Bayesian tracking and Kalman filtering framework on the three peaks closely follow the original peak latency. Although the signal is affected by noise, the tracking results still reflect the trend of the original position very well. We can conclude that the tracking algorithm effectively




Fig. 4 Tracking results on a simulated ICP dataset with 5% additive noise. Color curves represent the filtered peak latency on our simulated dataset using four different tracking models: Bayesian tracking, Kalman filter, LSTM, and MOCAIP

tracked the continuous waveform under this noise setting (i.e., 5%). The tracking result of LSTM is inconsistent with the groundtruth location, and its tracking result behaves differently for each peak and in different periods. The tracking result of p_1 is better than that of p_2 and p_3 . For p_3 , the detection is poor at the beginning and end of the tracking. The initialization phase of the LSTM could cause inconsistency in the initial part. Finally, the tracking result of MOCAIP captures the overall variations of the peak location but does not offer the same level of granularity as other techniques. MOCAIP is based on a clustering process to achieve peak detection. The input data are obtained by using a 1-min cluster average. However, it is worth mentioning that MOCAIP does not require a training process.

The tracking results in the presence of missing data are illustrated in Fig. 6, where the blue curve represents the waveform with missing segments, and the red represents the inferred output using one of the tracking algorithms. To enhance contrast, only tracking results for the missing data are displayed. The missing data segments are randomly distributed in the whole waveform range. In most cases, the tracking algorithm recovers the missing data by effectively capturing the trend of the data.

By observing the output predictions of the MOCAIP algorithm, we can see it can approximate the trend of the peak position even when data is missing. Although MOCAIP does not follow the details of the changes, it is still useful for getting an approximate estimation for missing data segments. On the other hand, LSTM provides a more refined estimate of the missing data. It should also be pointed out that only LSTM and MOCAIP algorithms




Fig. 5 Tracking results on a simulated ICP dataset with 5% additive noise. Color curves represent the filtered elevation of the first peak on our simulated dataset using four different tracking models: Bayesian tracking, Kalman filter, LSTM, MOCAIP




Fig. 6 LSTM and MOCAIP tracking results of peak latency are illustrated in the presence of missing data segments. Blue segments represent the peak latency detected on observed ICP waveforms, while the red segments represent the peak latency estimated in the presence of missing ICP waveforms

are used for missing data simulation because both Bayesian tracking and Kalman filter frameworks rely on the input for tracking, and the frameworks will not work when no input data are provided.

Discussion

Over the last two decades, machine learning algorithms have produced significant breakthroughs in various domains. In this study, we demonstrated the ability of several machine learning models to achieve high accuracy in a peak detection problem on a quasi-periodic signal, the intracranial pressure signal (ICP). Among the evaluated techniques, the peak detection error of kernel spectral regression (KSR) was the lowest, whether based on simulated or clinically acquired data.

We provided comparative results regarding tracking methods used to filter the peaks continuously. Bayesian inference, Kalman filtering, LSTM, and MOCAIP algorithms can represent the temporal dependence of neighboring pulses in the peak prediction process. The results of our experiments show that such frameworks are remarkably robust to noise and missing data. This could be explained by the fact that the temporal dependencies can play a significant role in maintaining the correct position of the peaks over time, as they are unlikely to change drastically between successive heartbeats.

ICP pulses arise from the blood pressure variation in the cerebral vasculature. In an ICP pulse, the specific distribution of sub-peaks is affected by capillary, arterial, and venous blood pressure pulses and their interactions with three major intracranial parts, including the brain tissue, the cerebral vasculature, and the cerebrospinal fluid circulatory system. Consequently, it is conceivable that ICP pulse morphological changes may provide reasonable indications of changes in these compartments. Also, these changes can be triggered by various pathological incidents, such as the narrowing cerebral arteries (vasospasm) after subarachnoid hemorrhage and the development of mass-occupying lesions after a brain injury. Therefore, the long-term continuous monitoring and recording of the ICP waveform provide the changing trend of the patient's physical condition, which is helpful for doctors to conduct pathological analysis of the state. Moreover, the tracking algorithm can predict the position of the ICP peak in a short period, which is also helpful for predicting the development of the disease in the clinical setting. In addition, given the interaction between biological signals, further study on the relationship between ICP and other biological signals to assist ICP waveform analysis is another direction to improve peak detection technology.

While we have made a special effort to identify techniques relevant to peak detection in ICP signals, the list of methods we have compared is not meant to be exhaustive. However, the set of experiments and the data can provide a baseline accuracy for developing and benchmarking future peak detection and tracking methods on ICP.

All the data and code used as part of our experiment will be made publicly available on the lab website of Prof. Scalzo (<http://www.fabiens.net>). To the best of our knowledge, this would become the first publicly available and curated dataset of ICP signals with both simulated and clinical sources. This is provided with the hope that the data and experimental protocol can serve as a benchmark for the development and evaluation of future peak detection methods in ICP.

Conclusion

This paper demonstrates that tracking of ICP waveform morphology can be performed in real-time with high accuracy using machine learning models such as kernel spectral regression (KSR), support vector machine (SVM), and LSTM. The acquisition of the ICP signal in a neuro-intensive care unit is often associated with signal loss and severe artifacts. To address these issues, our study demonstrated that peak detection models can be coupled with tracking models such as Kalman filter and nonparametric Bayesian inference to obtain robustness to temporary signal loss and improve the detection accuracy of the three landmarks. This paper also provides an ideal framework to benchmark future peak detection and tracking models. Although these tracking frameworks are demonstrated on ICP waveforms, they could, in principle, be used as part of the detection process of other quasi-periodic biological signals, such as ECG and CBFV.

Author contributions

M.W. and F.S. wrote the main manuscript text, and M.W., F.S., S.S., and S.K. prepared figures and a literature review. All authors reviewed the manuscript.

Funding

Prof. Miaomiao Wei was supported by the following grants: 1. Scientific Research Projects of Higher Education Institutions of Henan Province, China (23A510012). 2. Basic Scientific Research Foundation of Zhongyuan University of Technology, China (K2022QN020). 3. National Natural Science Foundation of China (62301624).

Availability of data and materials

Data and code to replicate experiments will be available on Prof. Fabien Scalzo's website <http://www.fabiens.net>.

Declarations

Ethics approval and consent to participate

The data were acquired at the Ronald Reagan Medical Center at the University of California, Los Angeles (UCLA), and the UCLA Internal Review Board (IRB) approved the usage of this archived dataset.

Competing interests

Not applicable.

Received: 8 May 2023 Accepted: 8 May 2024

Published online: 24 June 2024

References

- Cardoso ER, Rowan JO, Galbraith S. Analysis of the cerebrospinal fluid pulse wave in intracranial pressure. *J Neurosurg.* 1983;59(5):817–21.
- Cardoso ER, Reddy K, Bose D. Effect of subarachnoid hemorrhage on intracranial pulse waves in cats. *J Neurosurg.* 1988;69(5):712–8.
- Contant CF, Robertson CS, Crouch J, Gopinath SP, Narayan RK, Grossman RG. Intracranial pressure waveform indices in transient and refractory intracranial hypertension. *J Neurosci Methods.* 1995;57(1):15–25.
- Takizawa H, Gabra-Sanders T, Miller JD. Changes in the cerebrospinal fluid pulse wave spectrum associated with raised intracranial pressure. *Neurosurgery.* 1987;20(3):355–61.
- Portnoy HD, Chopp M. Cerebrospinal fluid pulse wave form analysis during hypercapnia and hypoxia. *Neurosurgery.* 1981;9(1):14–27.
- Chopp M, Portnoy HD. Systems analysis of intracranial pressure. comparison with volume-pressure test and csf-pulse amplitude analysis. *J Neurosurg.* 1980;53(4):516–27.
- Balestreri M, Czosnyka M, Steiner L, Schmidt E, Smielewski P, Matta B, Pickard J. Intracranial hypertension: what additional information can be derived from ICP waveform after head injury? *Acta Neurochir (Wien).* 2004;146(2):131–41.
- Czosnyka M, Guazzo E, Whitehouse M, Smielewski P, Czosnyka Z, Kirkpatrick P, Piechnik S, Pickard J. Significance of intracranial pressure waveform analysis after head injury. *Acta Neurochir (Wien).* 1996;138(5):531–41.
- Park C, Ryu SJ, Jeong BH, Lee SP, Hong C, Kim YB, Lee B. Real-time noninvasive intracranial state estimation using unscented kalman filter. *IEEE Trans Neural Syst Rehabil Eng.* 2019;27(9):1931–8.

10. Asgari S, Arevalo NK, Hamilton R, Hanchey D, Scalzo F. Cerebral blood flow velocity pulse onset detection using adaptive thresholding. In: 2017 IEEE EMBS International Conference on Biomedical Health Informatics (BHI), 2017; pp. 377–380.
11. Kim S, Hamilton R, Pineles S, Bergsneider M, Hu X. Noninvasive intracranial hypertension detection utilizing semisupervised learning. *IEEE Trans Biomed Eng.* 2013;60(4):1126–33.
12. Oh SL, Ng EY, Tan RS, Acharya UR. Automated diagnosis of arrhythmia using combination of CNN and ISTM techniques with variable length heart beats. *Comput Biol Med.* 2018;102:278–87.
13. Hochreiter S, Schmidhuber J. Long short-term memory. *Neural Comput.* 1997;9(8):1735–80. <https://doi.org/10.1162/neco.1997.9.8.1735>.
14. Hu X, Xu P, Scalzo F, Vespa P, Bergsneider M. Morphological clustering and analysis of continuous intracranial pressure. *IEEE Trans Biomed Eng.* 2009;56(3):696–705.
15. Hu X, Glenn T, Scalzo F, Bergsneider M, Sarkiss C, Martin N, Vespa P. Intracranial pressure pulse morphological features improved detection of decreased cerebral blood flow. *Physiol Measure.* 2010;31(5):679–95.
16. Jacobson AL. Auto-threshold peak detection in physiological signals. In: 2001 Conference Proceedings of the 23rd Annual International Conference of the IEEE Engineering in Medicine and Biology Society, 2001;3:2194–2195.
17. Kaur A, Agarwal R, Kumar A. Adaptive threshold method for peak detection of surface electromyography signal from around shoulder muscles. *J Appl Stat.* 2018;45(4):714–26. <https://doi.org/10.1080/02664763.2017.1293624>.
18. Slodzinski R, Hildebrand L, Vautz W. Peak detection algorithm based on second derivative properties for two dimensional ion mobility spectrometry signals. Berlin: Springer; 2013. p. 341–54. https://doi.org/10.1007/978-3-642-34471-8_28.
19. Kumar A, Ranganatham R, Komaragiri R, Kumar M. Efficient GRS complex detection algorithm based on fast Fourier transform. *Biomed Eng Lett.* 2019;9(1):145–51.
20. Rabbani H, Mahjoob MP, Farahabadi E, Farahabadi A. R peak detection in electrocardiogram signal based on an optimal combination of wavelet transform, Hilbert transform, and adaptive thresholding. *J Med Signals Sens.* 2011;1(2):91–8.
21. Chen H, Maharatna K. An automatic r and t peak detection method based on the combination of hierarchical clustering and discrete wavelet transform. *IEEE J Biomed Health Inform.* 2020;24(10):2825–32.
22. Jarman KH, Daly DS, Anderson KK, Wahl KL. A new approach to automated peak detection. *Chemometr Intell Lab Syst.* 2003;69(1):61–76.
23. Chanwimalueang T, von Rosenberg W, Mandic DP. Enabling r-peak detection in wearable ECG: combining matched filtering and hilbert transform. In: 2015 IEEE International Conference on Digital Signal Processing (DSP), 2015; pp. 134–138.
24. Nguyen T, Qin X, Dinh A, Bui F. Low resource complexity r-peak detection based on triangle template matching and moving average filter. *Sensors.* 2019;19(18):3997.
25. Sezan MI. A peak detection algorithm and its application to histogram-based image data reduction. *Comput Vis Graph Image Process.* 1990;49(1):36–51.
26. Halder B, Mitra S, Mitra M. Detection and identification of ecg waves by histogram approach. In: 2016 2nd International Conference on Control, Instrumentation, Energy Communication (CIEC), 2016; pp. 168–172.
27. Farashi S. A multiresolution time-dependent entropy method for GRS complex detection. *Biomed Signal Process Control.* 2016;24:63–71.
28. Harmer K, Howells G, Sheng W, Fairhurst M, Deravi F. A peak-trough detection algorithm based on momentum. In: 2008 Congress on Image and Signal Processing, vol. 4, 2008; pp. 454–458.
29. Deng H, Xiang B, Liao X, Xie S. A linear modulation-based stochastic resonance algorithm applied to the detection of weak chromatographic peaks. *Anal Bioanal Chem.* 2006;386(7–8):2199–205.
30. Panoulas KI, Hadjileontiadis LJ, Panas SM. Enhancement of r-wave detection in ecg data analysis using higher-order statistics. In: 2001 Conference Proceedings of the 23rd Annual International Conference of the IEEE Engineering in Medicine and Biology Society, vol. 1, 2001; pp. 344–347.
31. El Bouny L, Khalil M, Adib A. QRS complex detection based on smoothed nonlinear energy operator. In: 2018 9th International Symposium on Signal, Image, Video and Communications (ISIVC), 2018; pp. 191–196.
32. Hossain MB, Bashar SK, Walkey AJ, McManus DD, Chon KH. An accurate GRS complex and p wave detection in ECG signals using complete ensemble empirical mode decomposition with adaptive noise approach. *IEEE Access.* 2019;7:128 869–128 880.
33. El Bouny L, Khalil M, Adib A. R peak detection based on wavelet transform and nonlinear energy operator. In: Khoukhi F, Bahaj M, Ezziyani M, editors. *Smart data and computational intelligence*. Cham: Springer International Publishing; 2019. p. 104–12.
34. Dave T, Pandya U. R peak detection for wireless ECG using dwt and entropy of coefficients. *Int J Biomed Eng Technol.* 2020;34(3):268–83. <https://doi.org/10.1504/IJBET.2020.111472>.
35. Cai D, He X, Han J. SRDA: an efficient algorithm for large-scale discriminant analysis. *IEEE Trans Knowl Data Eng.* 2008;20(1):1–12.
36. Hasan MA, Reaz MBI, Ibrahimy MI. Fetal electrocardiogram extraction and r-peak detection for fetal heart rate monitoring using artificial neural network and correlation. In: *The International Joint Conference on Neural Networks*. 2011; 15–20.
37. Chang C-C, Lin C-J. LIBSVM: a library for support vector machines, 2001, software available at <http://www.csie.ntu.edu.tw/~cjlin/libsvm>.
38. Geurts P, Ernst D, Wehenkel L. Extremely randomized trees. *Mach Learn.* 2006;63(1):3–42.
39. Hagan M, Menhaj M. Training feedforward networks with the Marquardt algorithm. *IEEE Trans Neural Netw.* 1994;5(6):989–93.
40. Akhbari M, Niknazar M, Jutten C, Shamsollahi MB, Rivet B. Fetal electrocardiogram r-peak detection using robust tensor decomposition and extended Kalman filtering. *Comput Cardiol.* 2013;2013:189–92.

41. Scalzo F, Asgari S, Kim S, Bergsneider M, Hu X. Bayesian tracking of intracranial pressure signal morphology. *Artif Intell Med.* 2012;54(2):115–23. <https://doi.org/10.1016/j.artmed.2011.08.007>.
42. Welch G, Bishop G. An introduction to the Kalman filter. USA, Tech. Rep., 1995.
43. Sudderth EB, Ihler AT, Isard M, Freeman WT, Willsky AS. Nonparametric belief propagation. *Commun ACM.* 2010;53(10):95–103. <https://doi.org/10.1145/1831407.1831431>.
44. Hu X, Xu P, Lee D, Vespa P, Bergsneider M. An algorithm of extracting intracranial pressure latency relative to electrocardiogram r wave. *Physiol Meas.* 2008;29:459–71.
45. Afonso VX, Tompkins WJ, Nguyen TQ, Luo S. ECG beat detection using filter banks. *IEEE Trans Biomed Eng.* 1999;46(2):192–202.

Publisher's Note

Springer Nature remains neutral with regard to jurisdictional claims in published maps and institutional affiliations.