



www.bioinformation.net
Volume 20(8)

Research Article

Received August 1, 2024; Revised August 31, 2024; Accepted August 31, 2024, Published August 31, 2024

DOI: 10.6026/973206300200921

BIOINFORMATION 2022 Impact Factor (2023 release) is 1.9.

Declaration on Publication Ethics:

The author's state that they adhere with COPE guidelines on publishing ethics as described elsewhere at <https://publicationethics.org/>. The authors also undertake that they are not associated with any other third party (governmental or non-governmental agencies) linking with any form of unethical issues connecting to this publication. The authors also declare that they are not withholding any information that is misleading to the publisher in regard to this article.

Declaration on official E-mail:

The corresponding author declares that lifetime official e-mail from their institution is not available for all authors

License statement:

This is an Open Access article which permits unrestricted use, distribution, and reproduction in any medium, provided the original work is properly credited. This is distributed under the terms of the Creative Commons Attribution License

Comments from readers:

Articles published in BIOINFORMATION are open for relevant post publication comments and criticisms, which will be published immediately linking to the original article without open access charges. Comments should be concise, coherent and critical in less than 1000 words.

Disclaimer:

The views and opinions expressed are those of the author(s) and do not reflect the views or opinions of Bioinformation and (or) its publisher Biomedical Informatics. Biomedical Informatics remains neutral and allows authors to specify their address and affiliation details including territory where required. Bioinformation provides a platform for scholarly communication of data and information to create knowledge in the Biological/Biomedical domain.

Citation: Parmar

Edited by P Babaji & P Kanguane
Bioinformation 20(8): 921-925 (2024)

Dependability of Osstell ISQ's for measuring implant stability

Vijay Parmar¹, Nureldeen AN Elhammali^{2,*}, Omar Basheer Altaher Mohammed³, Meghna Chauhan⁴, Pushkar Gupta⁵, Abhigyan Manas⁶, Ankita Raj⁷ & Hrithik Chetani⁸

¹Department of Prosthodontics, Goenka Research Institute of Dental Science, Gandhinagar, Gujarat, India; ²Department of Oral and Maxillofacial Surgery, Faculty of Dentistry, Sirte University, Libya; ³Department of Oral and Maxillofacial Surgery, Sebha Dental Faculty, Sebha University, Libya; ⁴Department of Dentistry, World College of Medical Sciences & Research and Hospital, Jhajjar, Haryana, India; ⁵Department of Prosthodontics and Crown & Bridge, Hitkarini Dental College and Hospital, Jabalpur, MP, India; ⁶Department of Oral and Maxillofacial Surgery, Babu Banarsi Das College of Dental Sciences, Lucknow, UP, India; ⁷Department of Oral and Maxillofacial Surgery, Rama Dental College, Hospital and Research Centre, Kanpur, Uttar Pradesh, India; ⁸Department of Oral Medicine and Radiology, Kalinga Institute of Dental Sciences, KIIT Deemed to be University, Patia, Bhubaneswar-751024, Odisha, India; *Corresponding author

Affiliation URL:

<https://www.grids.co.in/>
<https://su.edu.ly/en/>
<https://sebhau.edu.ly/en>
<https://wcmsrh.com/>
<https://hdch.hitkarini.com/>
<https://bbdcods.edu.in/>
<https://www.ramauniversity.ac.in/>
<https://kids.kiit.ac.in/>

Author contacts:

Vijay Parmar -E-mail: vijayparmar8597@gmail.com; Phone +91 7600630478
 Nureldeen AN Elhammali -E-mail:nurdentnur@yahoo.com; Phone +218911361480
 Omar Basheer Altaher Mohammed -E-mail:omar1973den@gmail.com; Phone +00218912225008
 Meghna Chauhan -E-mail:drmeghna27@gmail.com; Phone +91 9538122867
 Pushkar Gupta -E-mail: drpushkar1@gmail.com; Phone: +91 9406538208
 Abhigyan Manas -E-mail: abhigyanmanas@gmail.com, Phone No: 993652882
 Ankita Raj -E-mail:rajdrankita@gmail.com; Phone +91 9415465531
 Hrithik Chetani -E-mail: hrithikchetami.ch06@gmail.com, Phone No: +91981419525

Abstract:

Determining the optimal loading schedule and measuring implant stability at different times are critical tasks. Numerous tools have been created to assess implant-bone stability as a sign of a well-treated implant. Thus, the objective of this cross-sectional study was to estimate the validity of the Osstell ISQ system for assessing implant stability. Osstell ISQ was used to complete implant stability registers for 60 implants across 18 patients. Two distinct SmartPegs (types I and II) were used to complete six measurements on each implant, or three measurements in a row with each transducer. In the 1st, 2nd, and 3rd measurements with SmartPegs I and II, the average ISQ was 71.36, 71.31, and 71.65, and 71.02, 71.58, and 71.76, respectively. For SmartPegs I and II, equivalent values or variations below three ISQ points were found in 46.3% and 58.6% of the cases, respectively. Both SmartPegs had an intraclass correlation coefficient of 0.96, and they also had repeatability and reproducibility of 0.96. An intra-class correlation coefficient analysis reveals nearly excellent repeatability and reproducibility for the RFA system Osstell ISQ. Measurements of Osstell ISQ have excellent repeatability.

Keywords: Dental implants, resonance frequency analysis, ISQ, implant stability, Osstell.

Background:

Dental implants are becoming more and more popular as a tooth replacement option these days. With survival and success rates comparable to those attained with conventional load, immediate load in implants has shown to be a predictable treatment [1]. The durability and osseointegration of dental implants are critical components of their success. Implants must osseointegrated with alveolar bone in order to operate and sustain masticatory pressure. Implant osseointegration is the process by which the titanium surface of the implant comes into close contact with the alveolar bone without the need for soft tissue to about it. There are two phases of osseointegration: primary and secondary. During the initial stage, mechanical interaction with cortical bone is the primary means of achieving implant stability. On the other hand, bone remodelling and regeneration are used to achieve implant stability in the secondary stage [2]. Primary implant stability upon implantation is influenced by a number of variables, including implant design, surgical technique, and bone density. Since it restricts micro-movements in the contact between the implant and the bone, primary implant stability is the key determinant of successful initial loading [3].

After implant implantation, initial stability has typically been evaluated based on bone quality and movement using the Lekholm & Zarb jaw quality scale [4]. One can forecast a dental implant's prognosis based on its stability. An implant's stability was defined as its capacity to withstand vertical, rotational, and horizontal stresses. This capacity was used as a proxy for osseointegration and effective healing [5]. Stability refers to an implant's ability to sustain load in the axial, lateral, and rotational directions. The quantity and quality of bone, surgical technique, and implant design all affect primary stability. During the first week following implantation, the primary stability is crucial to the stability of the implant. After that, it dramatically declines to negligible levels at around two weeks. The development of new bone and bone remodelling at the implant-bone contact are prerequisites for secondary stability. A biological mechanism forms the basis of the secondary stability [6, 7].

Implant stability measures (ISQ) are utilised as a predictive sign for potential implant failure and as an indirect indicator to establish the time range for practical implant loading in clinical practice [7]. The degree to which an implant resists deformation

in response to an applied force is known as its stiffness. The mechanical characteristics of the bone at the implant placement site and the degree, to which the fixture is engaged with the osseous tissue, as dictated by implant geometry and surgical technique, are the two key factors that affect primary stability [8]. Primary stability can be measured in a number of ways, some of which entail non-invasive quantitative analysis like damping capacity analysis (DCA) and resonance frequency analysis (RFA). The Osstell ISQ Mentor (Osstell, Göteborg, Sweden) is one of the RFA devices. It measures resonance frequency values using a smart-peg sensor paired with an implant fixture. The resonance frequency measurements are then translated into an arbitrary implant stability scale value known as the implant stability quotient (ISQ). Implant damping properties are measured by DCA systems according to contact time [5]. The implant stability test (IST) is conducted using percussion with the more current AnyCheck® system [9].

Insertion torque, percussion testing, and radiographic examination are non-invasive techniques for assessing implant stability. There are a few things to take into account even though radiographic examination is non-invasive and offers useful information regarding the state of the implant and surrounding bone. Images from radiography can get distorted [10]. Resonance frequency analysis, often known as the OTTL method, is a non-invasive clinical technique that was described by Meredith Sävedalen, Sweden-based Integration Diagnostics Ltd. has been designing Osstell devices since 1999. For implant stability testing, this device has undergone multiple generations in the last ten years [3,11]. A piezoelectric transducer is used to generate a sinusoidal signal at a certain frequency intended to cause the implant to vibrate, the non-invasive diagnostic method works. The instrument measures the implant's resistance to vibration, converting it into an ISQ value (implant stability quotient) on a scale of 0 to 100, where 100 represent the maximum implant stability [3]. RFA has emerged as one of the most popular methods for evaluating implant

two samples. SPSS 21.0 (SPSS, Chicago, IL) software package for Microsoft Windows was used to finish the data analysis.

Results:

Eighteen patients received a total of sixty implants, of which fifty-eight were put in the posterior maxillary region. The anterior maxillary area received the placement of the final two implants. There were 39 implants measuring 9 mm, 13 measuring 11 mm, and 8 measuring 12 mm in length. The total average scores that were achieved for the various groups were between 71.45 ISQ (SmartPegs II) and 71.44 (SmartPegs I). In the initial measurement, the average register obtained with SmartPeg I was 71.36 ISQ ± DT 6.008; in the subsequent measures, it was 71.31 ISQ ± DT 6.246 and 71.65 ISQ ± DT 6.176, respectively. However, the average register recorded with SmartPeg II was 71.02 ISQ ± DT 6.176 in its first measurement, and 71.58 ISQ ± DT 6.064 and 71.76 ISQ ± DT 6.021 in its second and third measurements, respectively. In terms of reliability analysis, **Table 1** shows that the intraclass correlation coefficient (ICC) for both SmartPegs I and II was 0.96, with a 95% confidence interval of 0.95 to 0.97. This indicates an almost perfect degree of concordance. Measurement discrepancies between SmartPegs I and II were displayed in **Table 2**. Variations in SmartPeg I measurements were examined and categorised based on the variations: 0 indicates the same value; 1-3, 4-5, and >5 indicate differences of more than five ISQ points.

As a result, 78% of the second transducer's register and 74.3% of the first transducer's register differed by three ISQ points or less. The differences between the six completed measures were then evaluated by comparing the Smart Peg I and II measurements. Based on an analysis of variance model with repeated or intrasubject measures, the intraclass correlation coefficient was utilised to assess concordance among the difference measurements done with both SmartPegs. Both SmartPegs have repeat abilities of 0.96 and Osstell ISQ reproducibility's of 0.96.

Table 1: The ISQ results' mean and standard deviation using both transducers

| Transducer | Measure | Mean ±SD |
|--------------|---------|--------------|
| Smart Peg I | 1st | 71.36 ±6.008 |
| | 2nd | 71.31 ±6.246 |
| | 3rd | 71.65 ±6.176 |
| Smart Peg II | 1st | 71.02 ±6.064 |
| | 2nd | 71.58 ±6.368 |
| | 3rd | 71.76 ±6.021 |

Table 2: Measurement variations between global (all six measurements using both SmartPegs) and SmartPegs I, II

| SMARTPEG I | | SMARTPEG II | | GLOBAL (BOTH SMARTPEGS) | |
|------------|----------|-------------|----------|-------------------------|----------|
| Difference | Frequenc | Difference | Frequenc | Difference | Frequenc |
| s | y % | s | y % | s | y % |
| 0 | 16 | 0 | 12 | 0 | 2 |
| 1-3 | 32 | 1-3 | 40 | 1-3 | 39 |
| 4-5 | 8 | 4-5 | 5 | 4-5 | 11 |
| >5 | 4 | >5 | 3 | >5 | 8 |
| Total | 60 | Total | 60 | Total | 60 |

Discussion:

Resonance Frequency Analysis (RFA) is a non-invasive intraoral technique intended to evaluate the interface between the implant

and bone, potentially offering clinical proof of implant stability. RFA makes it possible to monitor an implant through a series of stability assessments and to evaluate indirectly how osseous remodelling surrounding the implant affects secondary implant stability [3,11]. The current study used a cross-sectional analysis to evaluate the Osstell ISQ system's dependability for determining implant stability. Regarding repeatability in implant stability evaluation, we discovered that the Osstell system has very high dependability.

Using the Osstell® and AnyCheck® devices, Okuhama discovered the association between primary and secondary implant stability. They recommend AnyCheck ® as a helpful tool for figuring out implant stability, including primary and secondary [9]. According to Lachmann . Osstell ISQ and Periotest shown adequate reliability in predicting the implant stability [12]. According to El-Sawy 's conclusion, the Periotest (PTVs) and Osstell (ISQs) methods may both provide accurate evaluations of implant stability [13]. Using Periotest M (PTV), Anycheck, and Osstell Mentor (ISQ), Esposito . assessed the stability of the implant. They came to the conclusion that Anycheck has demonstrated its relative dependability in contrast to Periotest M and Osstell ISQ Mentor [14]. Osstell ISQ and Penguin RFA are only dependable when the implants are placed in stiff materials, according to Buyukguclu [15].The Anycheck device's dependability and the impact of the healing abutment diameter on the Anycheck values (implant stability test, IST) were investigated by Lee The Periotest M and Anycheck scores as well as the ISQ and IST showed a substantial association with one another. The Anycheck readings were unaffected by the healing abutment's diameter [5]. Lee claim that Anycheck®'s simplicity made it easier to operate and more accessible than the Osstell® Beacon+ [10]. Studies have shown that implant stability could be accurately measured using the Periotest and Osstell ISQ devices [5].The link between RFA and DCA device outcomes, which show the stability of the same implant, has been documented in a number of earlier investigations. A study conducted in vitro revealed a robust association between the outcomes obtained using RFA and DCA devices [16]. Additionally, after surgery and two months later, Pang et al. demonstrated a high correlation between the ISQs and PTVs [17]. Reynolds assessed the impact of several clinical features on the values generated by the devices throughout these three intervals. They came to the conclusion that values recorded between the two measurement devices at implant placement, implant exposure, and three years after installation exhibited a weak to moderate degree of correlation [8]. The implant treatment's outcome is dependent on the implant's insertion torque (IT) in the bone. Bone resorption is encouraged when the IT is elevated. IT optimisation is thought to be essential for a successful implant therapy outcome [9]. Using the AnyCheck® device, Al-Jamal . showed a substantial association between primary stability and IT in clinical practice [18]. Using resonant frequency analysis, Haseeb . assessed the insertion torque (IT) and implant stability of

two distinct implant macro geometries in various bone densities. When comparing the new implant macrogeometry to the conventional implant macro geometry, they discovered a lower IT value. No correlation was seen between IT and implant stability [19].The Osstell® is a well-established tool that has been the subject of numerous researches. However, in order to assess implant stability, it is necessary to remove the healing abutment and connect the smart peg, which may have drawbacks and inconveniences. Osstell Beacon®, which was just released, operates wirelessly. It still needs a smart peg, though, and the healing abutment needs to be put in and taken out. According to Esposito , there was 0.16 mm of bone resorption annually as a result of the healing abutment being removed three times after implantation until the superstructure attachment period [14]. We discovered that the Osstell system is reliable for assessing implant stability. Additional research is required to validate the results.

Conclusions:

The findings of this study suggest that, following statistical analysis using the intraclass correlation coefficient, the resonance frequency analysis system Osstell ISQ exhibits "near perfect" repeatability and reproducibility. Thus, it can be said that in terms of repeatability, Osstell system measurements are quite dependable. As a result, a single measurement can be enough.

References:

[1] Esposito M . 2007 22:893. [PMID: 18271370]

[2] Sennerby L & Roos J. 1998 11:408. [PMID: 9922733]

[3] Herrero-Climent M 2013 18:e877.[

[4] Lekholm U & Zarb GA. Chicago: Quintessence; 2011.p.199.

[5] Lee DH 2020 12:83. [https://doi.org/10.4047/jap.2020.12.2.83]

[6] AL-Juboori MJ 2015 9: 243. [PMID: 26312095]

[7] Huang H 202010:629. [https://doi.org/10.1016/j.jobcr.2020.07.004]

[8] Reynolds I . 2023 1:1139407. [Doi: 10.3389/fdmed.2023.1139407]

[9] Okuhama Y . 2022 22:1.[https://doi.org/10.1186/s12903-022-02320-0]

[10] Lee J 202419:e0290595. [https://doi.org/10.1371/journal.pone.0290595]

[11] Meredith N. 1998 11:491. [PMID: 9922740]

[12] Lachmann S . 2006 17:75. [PMID: 16441787]

[13] El-Sawy MK 202023:173. [10.21608/AJDSM.2020.26065.1039]

[14] Esposito M . 2017 10:57. [PMID: 28327695]

[15] Buyukguclu G . 2018 27: 429. [DOI: 10.1097/id.0000000000000766]

[16] Krafft T . 2012 27:318. [PMID: 22442770]

[17] Pang KM . 2014 25:616.[PMID: 23278670]

[18] Al-Jamal MFJ &Al-Jumaily HA. 2021 32:e171. [PMID: 33705063]

[19] Haseeb SA 2021 22:665. [PMID: 34393124]