RESEARCH

Optimizing stent retrievers for mechanical enhancement and in vitro testing in acute ischemic stroke models

Jae-Won Lee^{1†}, Han-Ki Kim^{1†}, JinWoo Kim² and Hyuk Choi^{1*}

† Jae-Won Lee and Han-Ki Kim contributed equally to this work.

*Correspondence: hyuk76@korea.ac.kr

¹ Department of Convergence Medicine, College of Medicine, Korea University, Seoul, South Korea

2 Department of Bionanosystem Engineering, Graduate School, Jeonbuk National University, Jeonju, South Korea

Abstract

Background: Acute ischemic stroke (AIS) remains a major cause of morbidity and mortality worldwide. Mechanical thrombectomy, especially with stent retrievers, offers a promising treatment, particularly for patients ineligible for intravenous tissue plasminogen activator (IV tPA) therapy. This study aimed to develop and evaluate novel stent retriever designs to enhance mechanical properties and vessel compatibility.

Results: We evaluated four stent designs using finite-element analysis (FEA) to assess their mechanical properties. Based on these evaluations, Stent D emerged as the optimal design due to its superior elasticity and adaptability. Comparative testing of Stent D against commercial stents, Solitaire FR and Trevo XP ProVue, revealed the following metrics: radial forces of 3.77 ± 0.01 N for Solitaire FR, 3.92 ± 0.08 N for Trevo XP Pro-Vue, and 4.10 ± 0.07 N for Stent D; flexibility measurements of 0.38 ± 0.11 N for Solitaire FR, 0.91 \pm 0.11 N for Trevo XP ProVue, and 0.59 \pm 0.05 N for Stent D; deployment forces of 0.37 ± 0.02 N for Solitaire FR, 0.42 ± 0.04 N for Trevo XP ProVue, and 0.32 ± 0.02 N for Stent D; and recapture forces of 0.38 ± 0.01 N for Solitaire FR, 0.45 ± 0.02 N for Trevo XP ProVue, and 0.35 ± 0.01 N for Stent D. Thrombus retrieval rates were 96.16% for Solitaire FR and 95.51% for Stent D.

Conclusions: These fndings demonstrate that Stent D performs comparably to com‑ mercial stents, highlighting its efective performance in AIS treatment. Stent D shows promise as a candidate for further clinical evaluation due to its superior mechanical properties and efective thrombus retrieval capabilities.

Keywords: Stent retriever, Acute ischemic stroke, Stent design, Finite-element analysis, Performance evaluation

Background

Acute ischemic stroke (AIS) remains a major cause of morbidity and mortality world-wide, prompting continuous efforts to develop effective treatment strategies [\[1](#page-10-0), [2](#page-10-1)]. While intravenous tissue plasminogen activator (IV tPA) therapy has demonstrated efficacy in thrombolysis within a limited time window, its application is constrained by factors such as treatment eligibility criteria and the risk of hemorrhagic complications [\[3](#page-10-2), [4\]](#page-10-3). In recent years, mechanical thrombectomy has emerged as a promising alternative

© The Author(s) 2024. **Open Access** This article is licensed under a Creative Commons Attribution 4.0 International License, which permits use, sharing, adaptation, distribution and reproduction in any medium or format, as long as you give appropriate credit to the original author(s) and the source, provide a link to the Creative Commons licence, and indicate if changes were made. The images or other third party material in this article are included in the article's Creative Commons licence, unless indicated otherwise in a credit line to the material. If material is not included in the article's Creative Commons licence and your intended use is not permitted by statutory regulation or exceeds the permitted use, you will need to obtain permission directly from the copyright holder. To view a copy of this licence, visit [http://](http://creativecommons.org/licenses/by/4.0/) [creativecommons.org/licenses/by/4.0/.](http://creativecommons.org/licenses/by/4.0/)

or adjunctive therapy for AIS, particularly for patients with large vessel occlusions or those ineligible for IV tPA [\[5](#page-10-4), [6\]](#page-10-5). Among the various mechanical thrombectomy techniques, stent retriever deployment has garnered signifcant attention due to its ability to achieve rapid and complete recanalization of occluded vessels [[7,](#page-10-6) [8\]](#page-10-7). However, current stent retriever designs, exemplifed by the Solitaire FR (Medtronic, USA) and Trevo XP ProVue (Stryker, USA) devices, possess inherent limitations. These devices, composed of nitinol alloy tubes and fabricated using laser-cutting techniques, typically feature a closed-cell design that may compromise fexibility and conformability within the vasculature compared to open-cell designs [\[9](#page-10-8), [10](#page-10-9)]. Given these challenges, there is a growing interest in developing novel stent retriever designs that optimize mechanical properties while enhancing vessel compatibility. This study aims to address this need by introducing four innovative stent retriever prototypes. Through comprehensive evaluation utilizing fnite-element analysis (FEA) and bench-top testing (Fig. [1\)](#page-1-0), we seek to identify the most promising design candidate that not only improves recanalization rates but also minimizes the risk of vascular injury and procedural complications. Following optimization, the performance of the stent was compared with that of the two products.

Results and discussion

Stent elasticity and adaptability using the computational solid mechanics

Stent elasticity and adaptability were assessed using Midas NFX software, employing computational solid mechanics. To validate the stent test results, the FEA model's elasticity and adaptability were evaluated through fat-type radial force tests and three-point bending tests, designed to simulate real-world conditions [[11,](#page-10-10) [12\]](#page-10-11). Figure [2](#page-2-0) presents the evaluation of the radial force following the application of the fattype radial force test. Compression was performed by positioning the stent between upper and lower plates and applying a load equivalent to 50% (2.5 mm) of the stent's radius. Radial forces for stent A, stent B, stent C, and stent D were 0.360, 0.184, 1.106, and 0.330 N, respectively (Fig. [2](#page-2-0)a). Stent C showed the highest expansion force;

Fig. 1 Experimental apparatus. **A** Laser tube cutting machine, Starcut tube fber (Rofn-Baasel Lasertech). **B** Radial force test device, RF-200 (Heinz Schade GmbH). **C** Micro-universal testing machine, T0-100S (Test One). **D** Radial force test jig. **E** Flexibility test jig. **F** Deployment and recapture test jig. **G** Cerebrovascular phantom. **H** Artificial thrombus

Fig. 2 FEA Results of stent elasticity and adaptability. **A** Flat type Radial force test of stents: Stent was placed in between the top and bottom plate and the top plate was compressed to give a load up to 50% (2.2 mm) of the radius of stent. The radial force of stent A, stent B, stent C, and stent D were 0.360, 0.184, 1.106, and 0.330 N, respectively. **B** 3-point bending test of stents: The result of 3-point bending test when giving 6.6-mm forced displacement to each stent models was 0.286, 0.268, 0.356, 0.098 N, respectively

However, such excessively high radial force stents can lead to severe adverse events more frequently, including vessel damage, restenosis, and even hemorrhages [[13,](#page-10-12) [14](#page-10-13)]. the reaction force values measured following the application of an external load using a three-point bending test. A displacement of 3.2 mm was applied to the top load applicator, and the reaction force of each stent model was measured. Reaction forces for stent A, stent B, stent C, and stent D were 0.242, 0.166, 0.296 N, and 0.226N, respectively (Fig. [2](#page-2-0)b). Likewise, Stent C showed the highest reaction force, and stent B, the lowest. Although Stent C showed higher reaction forces, indicating greater rigidity, this could potentially translate to lower fexibility and adaptability in tortuous vessel anatomies. Conversely, Stent B's lower reaction force in the bending test suggests it is more fexible, which is benefcial for navigating complex vascular pathways

Fig. 3 Comparative performance evaluation of structural features. **A** Radial force, **B** fexibility, **C** deployment force, and **D** recapture force. All tests were conducted with n=3 replicates

Table 1 Actual measurement values of comparative evaluation

	Solitare FR	Trevo XP ProVue	Stent D
Radial force (N)	3.77 ± 0.01	$3.92 + 0.08$	$4.10 + 0.07$
Flexibility (N)	0.38 ± 0.11	0.91 ± 0.11	$0.59 + 0.05$
Deployment force (N)	$0.37 + 0.02$	$0.42 + 0.04$	$0.32 + 0.02$
Recapture force (N)	$0.38 + 0.01$	$0.45 + 0.02$	$0.035 + 0.01$

without causing additional trauma or vessel injury. A balance between radial force and fexibility is necessary to minimize these risks while ensuring the stent can adequately support the vessel walls [[15,](#page-11-0) [16](#page-11-1)]. Stent D demonstrated a balanced combination of adequate elasticity and adaptability, which enhances safety and therapeutic outcomes in cerebrovascular applications.

Structural performance comparative evaluation

In Fig. [3](#page-3-0) and Table [1,](#page-3-1) to compare the structural performance of stents, we evaluated the most critical factors of radial force, fexibility, deployment force, and recapture force against commercially available products in Korea, including Solitaire FR and Trevo XP ProVue.

Radial force is crucial, because insufficient radial force can lead to deformation of the stent, preventing it from expanding the vessel properly and fully removing the thrombus, which can result in secondary occlusion [[14,](#page-10-13) [17\]](#page-11-2). Radial force measurements were conducted using Heinz's RF-200 (Heinz Schade GmbH, Mittelstadt, Germany) equipment to assess flat radial force. The stents were compressed by applying force from the top plate via a load applicator until they reached 50% of their diameter, and the reaction force was measured after achieving stent displacement. The results of the radial force examination indicated that Stent D exhibited the highest values: 3.77 ± 0.01 , 3.92 ± 0.08 , and 4.10 ± 0.07 N for Solitaire FR, Trevo XP ProVue, and Stent D, respectively, as depicted in Fig. [3](#page-3-0)a and Table [1.](#page-3-1) The high radial force of Stent D indicates its ability to maintain shape within the vessel, ensuring efective expansion and better apposition to

the vessel wall. This is critical for preventing thrombus migration and securely capturing and removing the thrombus.

Flexibility is critical for the stent's ability to navigate through tortuous and curved vascular pathways. If the stent lacks fexibility and exerts excessive force to revert to its original straight shape, it can cause vessel injury and subsequent infammatory responses, leading to restenosis [[17](#page-11-2)]. Flexibility tests were conducted using the three-point bending test according to ASTM F2606-08 standards, and force measurements corresponding to the bending displacement of the thrombectomy device were obtained using a Micro-Universal Testing Machine (UTM) system (T0-100S, Test One, Republic of Korea). The results in Fig. [3b](#page-3-0) and Table [1](#page-3-1) showed respective values of 0.38 ± 0.11 N for Solitare FR, 0.91 ± 0.11 N for Trevo XP ProVue, and 0.59 ± 0.05 N for Stent D. Since lower values indicate higher fexibility, this suggests better fexibility in the order of Solitaire FR, Stent D, and Trevo XP ProVue. It was observed that during the initial phase of the three-point bending test, there was a sharp increase in load, followed by a signifcant drop. This initial peak in force can be attributed to the contact between the stent and the testing apparatus, where the initial resistance to deformation is encountered. However, as the bending continued, the force required decreased substantially, likely due to the material properties of the nitinol stents, which exhibit superelasticity and nonlinear deformation characteristics [\[18](#page-11-3)–[20\]](#page-11-4). Accurately capturing these nuanced mechanical behaviors requires high-resolution load cells, capable of detecting subtle variations in force throughout the deformation process. Unfortunately, most UTM systems may not possess the necessary sensitivity to fully capture the low-force, high-deformation characteristics of nitinol stents, leading to potential underestimation of their performance in a clinical setting. This limitation should be considered when interpreting the results, as it may afect the precision of the measurements and the overall assessment of stent fexibility and resilience. Closed-cell stents, characterized by their interconnected struts, ofer greater radial support but are inherently less fexible compared to open-cell stents. Open-cell stents, with their segmented structures, provide enhanced axial fexibility, allowing them to better conform to the intricate and curved pathways of blood vessels [[21\]](#page-11-5). Stent D features a hybrid structure that combines both open and closed cells. This hybrid design efectively balances axial fexibility with radial stifness, ensuring that the stent can navigate through complex vessel anatomies without causing damage, thereby reducing the risk of post-procedural complications such as restenosis.

Deployment force is the force required to advance the stent to the target site. A low deployment force is crucial, because it helps the device penetrate the clot efectively and quickly upon deployment, improving the chances of successful clot removal with fewer attempts while protecting the blood vessel walls from damage $[22]$ $[22]$. The evaluation method for deployment force involved securing the catheter to the UTM and utilizing a compression module to push the thrombectomy device out of the catheter, measuring the force generated in this process. In Fig. [3c](#page-3-0) and Table [1](#page-3-1), the respective values were 0.37 ± 0.02 N for Solitaire FR, 0.42 ± 0.04 N for Trevo XP ProVue, and 0.32 ± 0.02 N for Stent D. Unsuccessful stent delivery to the target lesion has been associated with both micro and macroscopic damage to the stent structure. Stent D exhibited a moderate deployment force, ensuring ease of use during the procedure without compromising the integrity of the catheter or the device, thus facilitating smoother and safer stent delivery.

Recapture force is the force required to retract the thrombectomy device back into the catheter after it has been deployed, which is critical for repositioning or removing the device. High recapture force can hinder procedural efficiency and pose risks of the stent detaching from the catheter. For recapture force assessment, the catheter with the thrombectomy device deployed to approximately 50% was secured to the UTM, and a tension module was used to pull the thrombectomy device back into the catheter, meas-uring the force generated during this process. The results in Fig. [3](#page-3-0)d and Table [1](#page-3-1) showed respective values of 0.38 ± 0.01 , 0.45 ± 0.02 , and 0.035 ± 0.01 N. Stent D's intermediate recapture force ensures that the stent can be easily repositioned or retrieved without considerable difculty, thereby improving overall procedural safety and fexibility.

Tis study comparatively analyzed the performance of three stents using universally recognized performance testing methods. The performance of Stent D observed in the tests is crucial for achieving successful clot removal with minimal attempts, demonstrating the potential to reduce procedure time and minimize risks associated with vascular manipulation.

Thrombus retrieval rate evaluation

Following mechanical thrombectomy, patients with AIS may experience secondary occlusion of the cerebral artery due to incomplete thrombus removal, which can subsequently lead to occlusion of smaller vessels [\[23](#page-11-7)]. To evaluate the efectiveness of the stents in thrombus retrieval, we used AIS model with cerebrovascluar phantom and artificial clots made from polymer hydrogel. The experimental setup, shown in Fig. [4](#page-6-0), includes a cerebrovascular phantom capable of simulating the vascular conditions, an artifcial thrombus composed of 88.9% PAAM and 93% solvent (100 wt% DI water), a pulsatile flow pump (FlowTek125, United Biologics, Inc.), and a fluid reservoir. The pulsatile flow pump replicates physiological blood flow, allowing us to assess the stent's performance under conditions that mimic real blood flow dynamics. This setup enabled us to conduct three mock procedures to determine the efficiency of the stents in retrieving thrombi $[24, 25]$ $[24, 25]$ $[24, 25]$ $[24, 25]$. The thrombus retrieval rate was calculated as the percentage reduction in thrombus weight before and after the intervention using the following equation:

$$
Recovery\ rate\ (\%) = \frac{Pre-Introduction\ Clot\ weight\ (g)}{Post-International\ Clot\ weight\ (g)} \times 100
$$

The results for In vitro blood clot removal rate measurement are presented in Table [2](#page-6-1). The comparison of blood clot removal rates between Solitare FR and Stent D revealed a recovery rate of 96.16% for Solitare FR and 95.51% for Stent D. Both stents achieved thrombus retrieval rates of over 90% when inserted into the vascular phantom model. These findings suggest that there is no significant difference in thrombus removal performance between the two stents, indicating comparable results.

Conclusion

In this study, we proposed various shapes and designs of stent retriever for application in AIS patient models and determined the optimal stent design through FEA. After thorough analysis, Stent D emerged as the optimized design choice. This newly developed

Fig. 4 Schematic and experimental setup for thrombus retrieval rate evaluation using a cerebrovascular phantom model. Schematic representation (**B**) and detailed visualization (**A**) of the experimental setup: (1) cerebrovascular phantom capable of thrombus loading (**C**). (2) Artifcial thrombus composed of 88.9% PAAM and 93% solvent (DI water, 100 wt%). (3) Pulsatile fow pump, FlowTek125 (United Biologics, Inc.), and (4) fuid reservoir

Table 2 In vitro blood clot removal rate measurement results

stent exhibits promising mechanical properties and structural performance, comparable to, and in some aspects superior to, existing commercially available products such as Solitaire FR and Trevo XP ProVue. Its well-balanced combination of mechanical attributes, including adequate radial force, optimal fexibility, moderate deployment force, and robust recapture force, positions it as a suitable candidate for further clinical evaluation. Moreover, the thrombus retrieval rate evaluation conducted using the vascular phantom model yielded promising results, indicating that Stent D efectively retrieves thrombi in AIS patient models. Tis underscores the comprehensive functionality of Stent D, not only in terms of mechanical performance but also in its ability to address the primary objective of thrombus removal.

However, despite the positive outcomes from this study, it has limitations. Since stents are implanted within the human body, mechanical properties alone are not sufficient considerations. A limitation of this study is the absence of biocompatibility evaluation. Therefore, in future studies, we plan to conduct biocompatibility tests in accordance with ISO 10993 standards to ensure the cytocompatibility of our stent design. Specifcally, we will investigate methods such as electropolishing and surface coatings with materials like oxide flms and hyaluronic acid to improve the biocompatibility of the nitinol stents. Tese future investigations will provide a comprehensive understanding of Stent D's safety and efectiveness in vivo, paving the way for larger-scale clinical trials and extended follow-up studies that are crucial to fully establish its clinical efficacy and safety in thrombectomy procedures.

Methods

Model geometry

Tree-dimensional (3D) computational models were designed using CATIA P3 V5R18 (Dassault Systems, Vélizy-Villacoublay, France) software resulting in four types of stents, namely stent A, stent B, stent C and stent D. Nitinol tubes for stent manufacturing was purchased from Vascotube GmbH (Germany) with 5.0 mm of outer diameter and 0.08 mm of wall thickness. Stents were designed to prevent intraluminal stent protrusion, and the strut was manufactured with a thickness, width, length of 0.08, 0.075, and 0.15 mm, respectively, containing eight bands. Our stent design was based on the well-known clinically effective Enterprise stent. Figure [5](#page-8-0) shows the four stent models designed. Stent A was designed with alternating square and rectangular arrangements, utilizing S-shaped linkers (Fig. [5a](#page-8-0)). Stent B modifed the typical diamond strut structure into an S-shape (Fig. [5b](#page-8-0)). Stent C arranged leaf-like shapes in a zig-zag pattern, connected by straight peak-to-valley linkers (Fig. [5c](#page-8-0)). Stent D featured a hybrid cell structure by adding open cells to the design of Stent A (Fig. [5d](#page-8-0)).

Computational solid mechanics

The elasticity and adaptability of the stents were assessed using Midas NFX software (Midas IT, London, UK), leveraging computational solid mechanics. The nonlinear static mode was chosen for the simulation to accommodate the large deformations and superelasticity exhibited by the nitinol stent during fat-type and three-point bending tests. Material properties for the nitinol stents were defned with an elastic modulus of 28 GPa, a Poisson's ratio of 0.3, a material density of 6.45 g/cc, and a tensile strength of 100 MPa,

Fig. 5 Model geometry of stent: A stent A, **B** stent B, **C** stent C, and **D** stent D. Strut was manufactured with thickness of 0.08 mm, width of 0.075 mm, length of 0.15 mm and 8 bands

while the jigs were modeled as rigid bodies. Stents and jigs were both meshed with hexahedral solid fnite elements, with element sizes set to 0.04 mm for the stents and 0.1 mm for the jigs. To validate these assessments, fat-type radial and three-point bending tests were performed on the FEA model to replicate real-world conditions. Boundary conditions were applied by fxing the bottom plate, while the top plate was constrained in all directions except for displacement along the *y* axis, where displacement loading was allowed. During the high elasticity evaluation, stents were compressed up to 50% of their diameter by enforcing displacement through a load applicator on the top plate, with the reaction force measured after achieving a 2.5 mm displacement. Adaptability was evaluated by measuring the reaction force during the three-point bending test, where enforced displacement was applied to replicate the experimental setup. The reaction forces were recorded as the upper load applicator induced defection in the stent.

Manufacturing of stents

For the manufacturing of NiTi alloy stents, the standard fabrication process included laser machining, chemical treatment, and heat treatment. The NiTi alloy tubes (Johnson Matthey Inc. in San Jose, CA, USA) was machined using a photolithography-based laser processing method (Starcut tube fber, Rofn Bassel Lasertech, Germany) with a power output of 55 W, a laser wavelength of 16.5 µm (pulse), and an oxygen cooling system. Following laser cutting, chemical etching was performed, which involved two steps: cleaning and etching. The primary goal of chemical polishing was to eliminate impurities, defects, residual stress, and oxide films from earlier operations. The cleaning stage, lasting 30 min at room temperature, focused on removing oxides and organic materials. In addition, a deoxidizing process lasting 30 s at room temperature was conducted to remove oxide layers. Organic residues on the nitinol stent surfaces were removed using commercial solutions, MicroClean Ti solution™ (RDZ-1849; RD Chemical, Mountain

View CA, USA), followed by MicroClean MV™ (RDZ-1995; RD Chemical, Mountain View CA, USA), applied at room temperature for 30 s to eliminate the oxide flm. Heat treatment was performed using a high-vacuum annealing furnace (HT-1000, Heinz Schade, Germany) to produce a stable oxide layer with the desired mechanical properties for the nitinol stents. The heat treatment parameters were selected within the range of 450–600 °C and 10–30 min [\[26](#page-11-10), [27](#page-11-11)]. In this study, we chose a heat treatment condition of 510 °C for 10 min, as it resulted in the most stable oxide layers and suitable mechanical properties for the nitinol stents.

Cerebrovascular thrombus phantom model

Computed tomography data from patients with cerebrovascular diseases were collected to fabricate a silicone phantom model mimicking actual brain vascular vessels. Transparent silicone was selected as the vascular material, with an inner diameter of 3.5 mm and a wall thickness ranging from 0.25 to 0.5 mm. The curved segment of the model measured 15.88 mm in length, with a radius (*R* value) of 3.57 mm. Access ports were created in the phantom model at the middle cerebral artery, carotid artery, and vertebral artery to facilitate fluid flow and blood clot loading. The working fluid was composed of a blend of water (60% by weight) and glycerol (40% by weight) to achieve a density and dynamic viscosity that closely approximate those of blood $(1.09\pm0.03 \text{ g/ml}$ and $3.98\pm0.14 \text{ cP}$, respectively) at an operating temperature of 22.2 $^{\circ}$ C [[28](#page-11-12)]. The experiments were conducted with a steady inlet fow rate, ensuring optimal conditions for simulating physiological conditions. To confrm the integrity of the phantom model, we assessed the device's wall apposition using commercially available central circulation thrombectomy catheters. The evaluation ensured that the thrombectomy device maintained consistent contact with the vessel wall, efectively adapting to its curvature and demonstrating excellent wall apposition while retaining a cylindrical shape.

Preparation of thrombus

For the in-vitro evaluation of thrombus removal performance, we synthesized a polymer hydrogel to create an artifcial thrombus model. We synthesized thrombi based on polymer hydrogels with two diferent properties: hard and soft types, following the method applied in previous studies [\[29](#page-11-13), [30\]](#page-11-14). In brief, Polyacrylamide and alginate (PAAM-Alg) with a crosslinking agent were utilized. The composition for hard thrombus consisted of 88.9% PAAM and 85% solvent (100 wt% DI water), while soft thrombus comprised 88.9% PAAM and 93% solvent (100 wt% DI water). In this study, the hard clot frequently fractured during the testing process, whereas the soft clot exhibited physical properties more similar to actual thrombi, thereby facilitating the evaluation process. For the in vitro clot removal test, the synthesized clot model was cut to a diameter of 3 mm and a length of 20 mm for injection into the vascular model. The clot was inserted into the vascular model through an 8F guide catheter (Terumo, Tokyo, Japan) placed proximally at the target site, which had a diameter of 2–3 mm. The embedding time of the clot was set at 15 min, allowing it to settle in the target site, and the stent retriever was deployed across the clot through a 0.021-in microcatheter (Prowler Select Plus, Johnson & Johnson Med-Tech, USA). After 5 min of embedding time, the stent retriever and microcatheter were

retrieved together into the guide catheter. The thrombus retrieval rate was evaluated as the percentage of pre-intervention to post-intervention clot weight.

Acknowledgements

Not applicable.

Author contributions

Jae-Won Lee, Han-Ki Kim, Hyuk Choi conceived and designed the experiments. Han-Ki Kim and JinWoo Kim performed the experiments and Data collection. Jae-Won Lee contributed to the writing of the manuscript and analysed the data. Hyuk Choi performed the Critical revision of article. All authors reviewed the manuscript.

Funding

This work was supported by the National Research Foundation of Korea (NRF) grant funded by the Korea government (2020R1F1A1068910).

Availability of data and materials

The datasets generated and/or analysed during the current study are not publicly available due [REASON WHY DATA ARE NOT PUBLIC] but are available from the corresponding author on reasonable request.

Declarations

Ethics approval and consent to participate Not applicable.

Consent for publication

Not applicable.

Competing interests

The authors declare no competing interests.

Received: 8 August 2024 Accepted: 5 November 2024 Published online: 27 November 2024

References

- 1. Yuki I, Kan I, Vinters H, Kim R, Golshan A, Vinuela F, et al. The impact of thromboemboli histology on the performance of a mechanical thrombectomy device. Am J Neuroradiol. 2012;33(4):643–8.
- 2. Feigin VL, Forouzanfar MH, Krishnamurthi R, Mensah GA, Connor M, Bennett DA, et al. Global and regional burden of stroke during 1990–2010: fndings from the Global Burden of Disease Study 2010. Lancet. 2014;383(9913):245–54.
- 3. Hacke W, Kaste M, Bluhmki E, Brozman M, Dávalos A, Guidetti D, et al. Thrombolysis with alteplase 3 to 4.5 hours after acute ischemic stroke. N Engl J Med. 2008;359(13):1317–29.
- 4. Emberson J, Lees KR, Lyden P, Blackwell L, Albers G, Bluhmki E, et al. Effect of treatment delay, age, and stroke severity on the efects of intravenous thrombolysis with alteplase for acute ischaemic stroke: a meta-analysis of individual patient data from randomised trials. Lancet. 2014;384(9958):1929–35.
- 5. Saver JL, Goyal M, Bonafe A, Diener HC, Levy EI, Pereira VM, et al. Stent-retriever thrombectomy after intravenous t-PA vs t-PA alone in stroke. N Engl J Med. 2015;372(24):2285–95.
- 6. Goyal M, Menon BK, van Zwam WH, Dippel DW, Mitchell PJ, Demchuk AM, et al. Endovascular thrombectomy after large-vessel ischaemic stroke: a meta-analysis of individual patient data from fve randomised trials. Lancet. 2016;387(10029):1723–31.
- 7. Li J, Tiberi R, Bhogal P, Buhk J-H, Behme D, Tomasello A, et al. Impact of stent-retriever tip design on distal embolization during mechanical thrombectomy: a randomized in vitro evaluation. J NeuroInterv Surg. 2024;16(3):285-9.
- 8. Nogueira RG, Pinheiro A, Brinjikii W, Abbasi M, Al-Bayati AR, Mohammaden MH, et al. Clot composition and recanalization outcomes in mechanical thrombectomy. J Neurointerv Surg. 2024;16(5):466–70.
- 9. Mueller-Kronast NH, Zaidat OO, Froehler MT, Jahan R, Aziz-Sultan MA, Klucznik RP, et al. Systematic evaluation of patients treated with neurothrombectomy devices for acute ischemic stroke: primary results of the STRATIS registry. Stroke. 2017;48(10):2760–8.
- 10. Alawieh A, Vargas J, Fargen KM, Langley EF, Starke RM, De Leacy R, et al. Impact of procedure time on outcomes of thrombectomy for stroke. J Am Coll Cardiol. 2019;73(8):879–90.
- 11. Migliavacca F, Petrini L, Colombo M, Auricchio F, Pietrabissa R. Mechanical behavior of coronary stents investigated through the fnite element method. J Biomech. 2002;35(6):803–11.
- 12. Song K, Bi Y, Zhao H, Wu T, Xu F, Zhao G. Structural optimization and fnite element analysis of poly-l-lactide acid coronary stent with improved radial strength and acute recoil rate. J Biomed Mater Res B Appl Biomater. 2020;108(7):2754–64.
- 13. Shobayashi Y, Tateshima S, Tanishita K. Load-dispersing design with twined-spring geometry of a distensible intracranial stent for cerebral aneurysms. J Biorheol. 2013;26:29–37.
- 14. Katz JM, Hakoun AM, Dehdashti AR, Chebl AB, Janardhan V, Janardhan V. Understanding the radial force of stroke thrombectomy devices to minimize vessel wall injury: mechanical bench testing of the radial force generated by a novel braided thrombectomy assist device compared to laser-cut stent retrievers in simulated MCA vessel diameters. Interv Neurol. 2020;8(2–6):206–14.
- 15. Snowhill P, Nosher J, Silver F. The importance of radial forces in vascular stent design. Biomaterials engineering and devices: human applications: Volume 1 Fundamentals and vascular and carrier applications: Springer; 2000. p. 145–60.
- 16. Ishihara R. Endoscopic stenting for malignant dysphagia in patients with esophageal cancer. Curr Oncol. 2023;30(7):5984–94.
- 17. Pan C, Han Y, Lu J. Structural design of vascular stents: a review. Micromachines. 2021;12(7):770.
- 18. Duerig T, Pelton A, Stöckel D. An overview of nitinol medical applications. Mater Sci Eng, A. 1999;273:149–60.
- 19. Berti F, Petrini L, Spagnoli A. A discussion about multi-axial fatigue criteria for NiTinol cardiovascular devices. Procedia Struct Integrity. 2018;13:813–8.
- 20. Kossa A, McMeeking RM. Bending of a Nitinol cantilever and its fatigue performance. Extreme Mech Lett. 2021;42: 101083.
- 21. Ahmed SU, Chen X, Peeling L, Kelly ME. Stentrievers: an engineering review. Interv Neuroradiol. 2023;29(2):125–33.
- 22. Prince HC, Saliba AJ, Wheeler J, Bruder S. Development of the Trevo ProVue Retriever for intracranial clot removal in acute ischemic stroke. Ann N Y Acad Sci. 2014;1329(1):107–15.
- 23. Pfaf J, Herweh C, Pham M, Schieber S, Ringleb P, Bendszus M, et al. Mechanical thrombectomy of distal occlusions in the anterior cerebral artery: recanalization rates, periprocedural complications, and clinical outcome. Am J Neuroradiol. 2016;37(4):673–8.
- 24. Madjidyar J, Nerkada L, Larsen N, Wodarg F, Hensler J, Jansen O, editors. Choosing an efective and safe direct aspiration setup for tortuous anatomy in acute ischemic stroke: in vitro study in a physiological flow model. RöFo-Fortschritte auf dem Gebiet der Röntgenstrahlen und der bildgebenden Verfahren; 2021: Georg Thieme Verlag KG.
- 25. Sanchez S, Cortiñas I, Villanova H, Rios A, Galve I, Andersson T, et al. ANCD thrombectomy device: in vitro evaluation. J NeuroInterv Surg. 2020;12(1):77–81.
- 26. Vojtěch D, Voděrová M, Fojt J, Novák P, Kubásek T. Surface structure and corrosion resistance of short-time heattreated NiTi shape memory alloy. Appl Surf Sci. 2010;257(5):1573–82.
- 27. Elahinia MH, Hashemi M, Tabesh M, Bhaduri SB. Manufacturing and processing of NiTi implants: a review. Prog Mater Sci. 2012;57(5):911–46.
- 28. Riley J, Price N, Saaid H, Good B, Aycock K, Craven B, et al. In vitro clot trapping efficiency of the FDA generic inferior vena cava flter in an anatomical model: an experimental fuid–structure interaction benchmark. Cardiovasc Eng Technol. 2021;12:339–52.
- 29. Sun J-Y, Zhao X, Illeperuma WR, Chaudhuri O, Oh KH, Mooney DJ, et al. Highly stretchable and tough hydrogels. Nature. 2012;489(7414):133–6.
- 30. Merritt W, Holter AM, Beahm S, Gonzalez C, Becker TA, Tabor A, et al. Quantifying the mechanical and histological properties of thrombus analog made from human blood for the creation of synthetic thrombus for thrombectomy device testing. J Neurointerv Surg. 2018;10(12):1168–73.

Publisher's Note

Springer Nature remains neutral with regard to jurisdictional claims in published maps and institutional afliations.