



OPEN

Accuracy of zero-heat-flux thermometry and bladder temperature measurement in critically ill patients

Anselm Bräuer¹✉, Albulena Fazliu¹, Thorsten Perl², Daniel Heise¹, Konrad Meissner¹ & Ivo Florian Brandes¹

Core temperature (T_{Core}) monitoring is essential in intensive care medicine. Bladder temperature is the standard of care in many institutions, but not possible in all patients. We therefore compared core temperature measured with a zero-heat flux thermometer (T_{ZHF}) and with a bladder catheter (T_{Bladder}) against blood temperature (T_{Blood}) as a gold standard in 50 critically ill patients in a prospective, observational study. Every 30 min T_{Blood} , T_{Bladder} and T_{ZHF} were documented simultaneously. Bland–Altman statistics were used for interpretation. 7018 pairs of measurements for the comparison of T_{Blood} with T_{ZHF} and 7265 pairs of measurements for the comparison of T_{Blood} with T_{Bladder} could be used. T_{Bladder} represented T_{Blood} more accurate than T_{ZHF} . In the Bland Altman analyses the bias was smaller (0.05 °C vs. – 0.12 °C) and limits of agreement were narrower (0.64 °C to – 0.54 °C vs. 0.51 °C to – 0.76 °C), but not in clinically meaningful amounts. In conclusion the results for zero-heat-flux and bladder temperatures were virtually identical within about a tenth of a degree, although T_{ZHF} tended to underestimate T_{Blood} . Therefore, either is suitable for clinical use.

German Clinical Trials Register, DRKS00015482, Registered on 20th September 2018, <http://apps.who.int/trialsearch/Trial2.aspx?TrialID=DRKS00015482>.

Core temperature monitoring is one of the essential monitoring modalities in intensive care medicine and temperatures below or above the normal core temperature can be observed frequently. Hypothermia at admission of surgical patients is associated with many adverse outcomes like coagulopathy¹, increased bleeding², and higher transfusion rates³ as well as increased surgical site infections^{4,5} and in some studies even with mortality^{3,6–9}. However, fever is observed much more frequently in critically ill patients and warrants a diagnostic workup to determine the presence of potential infections¹⁰. The ideal temperature measurement method should provide reliable, reproducible values safely and conveniently¹⁰. Additionally, the device should be small, easy to use, comfortable, fast, continuous, noninvasive, low energy consuming and affordable¹¹. In many institutions bladder temperature (T_{Bladder}) is the standard of care because it is accurate, provides continuous readings, stable measurements regardless of urine flow rate and stays securely in place even during positioning of the patient¹⁰. Normally T_{Bladder} monitoring adds no additional invasiveness to the standard monitoring, because bladder catheters are nearly always used in critically ill patients.

However, not every patient meets the criteria for a bladder catheter. Patients with acute or chronic renal failure with anuria, after cystectomy or awake patients do not need a bladder catheter. With no indication for using a bladder catheter the use is associated with an unnecessary risk of nosocomial infection. Further on, patients that need irrigation of the bladder because of bleeding will not be appropriate for measurement of T_{Bladder} . In these situation an alternative approach is needed. Especially in alert patients a non-invasive monitor is helpful. Unfortunately, many of these methods are not very reliable¹². A better alternative might be a zero-heat flux (ZHF) thermometer that has been proven to be reliable in surgical patients^{13–16}. In general, zero-heat-flux thermometers consist of a thermal insulator adjacent to the skin that is covered by a servo-controlled electric heater. The heater is used to eliminate the flow of heat through the insulator, so that the temperature of the heater and skin temperatures are equal¹³. A validation study in critically ill patients is important because in patients undergoing surgery the most common thermal problem is perioperative hypothermia whereas in critically ill patients it is

¹Department of Anesthesiology, University Medical Center Göttingen, Robert-Koch Strasse 40, 37099 Göttingen, Germany. ²Department of General, Visceral and Pediatric Surgery, University Medical Center Göttingen, Göttingen, Germany. ✉email: abraeue@gwdg.de

fever. However, to date no large comparison in critically ill patients has been performed with an accepted gold standard like blood temperature. There is only one study available that has included a small number of patients with blood temperature as a reference method²⁶.

The aim of this study was to compare T_{Core} measured with a ZHF-thermometer (T_{ZHF}) and with T_{Bladder} against a gold standard T_{Blood} measured in the iliac artery or pulmonary artery to determine if the new ZHF-thermometer is more accurate than T_{Bladder} .

Methods

The current prospective clinical study was conducted in accordance with the declaration of Helsinki at the University Hospital of Göttingen, Germany, after obtaining local ethics committee approval (Ethics committee of the University Medicine Göttingen, Application number: 13/05/18) for the experimental protocol and registration on German Clinical Trials Register (DRKS00015482). According to the approval of the local ethical committee we used deferred (proxy) consent in emergency critical care research¹⁷ as the study was totally non-invasive and observational. If patients were able to give informed written consent this consent was used. If informed proxy consent was necessary, it was given in written form of the proxy. We did not exclude patients who did not recover and died during their hospital stay. The local ethics committee had approved this procedure. The article adheres to the STROBE guidelines¹⁸.

Critically ill adult patients already having a bladder catheter with a temperature probe and an arterial catheter with a temperature probe placed in the iliac artery (Pulsiocath Arterial Thermomodulation Catheter 5F; Pulsion Medical Systems AG, Munich, Germany) or a pulmonary artery catheter (Arrow Hands-Off Thermomodulation catheter 7F; Arrow International, Athlone, Ireland) in place were included in this study. The only exclusion criteria were pregnancy and refusal to take part in the study.

In all patients, core temperature was measured additionally with a single use, continuous, non-invasive ZHF-thermometer (3 M SpotOn Temperature Monitoring System, 3 M, St. Paul, MN, USA) attached to the lateral forehead of the patients.

Then every 30 min T_{Core} measured by T_{Blood} , T_{Bladder} and the ZHF-thermometer were documented at the same time points until the patient lost the T_{Blood} sensor or T_{Bladder} sensor, left the ICU or at least after 5 days. If data of a temperature source were missing the couple of data was not used for comparison. In addition to the temperature data age, weight, height, sex and medical diagnosis at admission to the Intensive Care Unit (ICU) were documented.

As a primary statistical method Bland–Altman statistics were used for interpretation of accuracy (bias = mean difference between methods) and precision (limits of agreement = 1.96 standard deviation) using the Bland and Altman random effects method for repeated measures data adjusted for unequal numbers of measurements per patient¹⁹. Additionally, we calculated the proportion of all differences that were within $\pm 0.5^\circ\text{C}$ or $\pm 1^\circ\text{C}$ of T_{Blood} .

For each of the two measurement modalities sensitivity, specificity, positive and negative predictive values for the detection of hypothermia and fever were calculated. Hypothermia was defined as a $T_{\text{Blood}} < 36^\circ\text{C}$ and fever was defined as $T_{\text{Blood}} > 38.3^\circ\text{C}$ ¹⁰.

Additionally, we performed an error grid analysis²⁰ to determine if some measurement differences would lead to wrong clinical decisions. The Zones were defined as follows:

Zone A begins with an area of a $\pm 0.5^\circ\text{C}$ error on either side of a perfectly accurate measurement between T_{Blood} and the temperature measured by T_{ZHF} or T_{Bladder} . Measurement errors smaller than $\pm 0.5^\circ\text{C}$ are considered by most authors as clinically not relevant. In addition, if both measurements indicate hypothermia $< 36^\circ\text{C}$ or fever $> 38.3^\circ\text{C}$ the absolute error is considered to be clinically irrelevant because the same treatment or diagnostic workup will be initiated.

Zone B describes the zone where measurement errors are $> 0.5^\circ\text{C}$ but this will not result in a clinical wrong decision. E.g. if T_{Blood} is 36.5°C and T_{ZHF} shows a temperature of 37.4°C both temperatures will not lead to active warming therapy or a diagnostic workup for infection.

In contrast Zone C indicates errors larger than 0.5°C that will lead to wrong clinical decisions and may do harm to the patient. e.g. if T_{Blood} is 34°C and T_{ZHF} shows 37°C the patient will not receive active warming although this would be indicated.

Results

55 potentially eligible patients were screened. Three patients could not be enrolled because we could not obtain proxy consent and two patients were not included due to technical problems. The remaining 50 patients were enrolled. 36 patients (72%) were male, 14 (28%) were female. Mean age was $61.9 (\pm 16.8)$ years, mean height was $1.75 (\pm 0.07)$ m, mean weight was $86.4 (\pm 36.3)$ kg resulting in a mean body mass index of $28.2 (\pm 11.3)$ kg/m². Of these patients 16 were suffering from sepsis, 18 patients had neurologic injury (subarachnoid hemorrhage, intracerebral hemorrhage), 6 patients had trauma, 4 patients had respiratory failure, 2 patients had accidental hypothermia, 3 patients had cardiac surgery, and 1 patient had visceral surgery. Of all 50 patients 49 had an arterial catheter with a temperature probe placed in the iliac artery and one patient had a pulmonary artery catheter with temperature probe. No patient was excluded from the study after enrolment.

Globally 3970.5 h were recorded. 7665 T_{Blood} values, 7086 values of T_{ZHF} and 7358 T_{Bladder} values were documented. 276 T_{Blood} values, 855 values of the ZHF-thermometer and 583 T_{Bladder} values were missing. The major reason for missing values was a disconnection of the temperature probes for transportation of the patient to the CT, OR, neuroradiology suite, or to the cardiac catheter lab. After these procedures the devices were often not reconnected immediately. Only 17 temperature values of T_{Bladder} and 16 values of T_{ZHF} were missing due to technical problems. 12 values below 30°C could not be recorded by the ZHF-thermometer because the device did not give a reading at these low temperatures.




Figure 1. Bland–Altman analysis for the zero-heat flux thermometer (T_{ZHF}) versus blood temperature (T_{Blood}).




Figure 2. Bland–Altman analysis for bladder temperature (T_{Bladder}) versus blood temperature (T_{Blood}).

This resulted in 7018 pairs of measurements for the comparison of T_{Blood} with T_{ZHF} and 7265 pairs of measurements for the comparison of T_{Blood} with T_{Bladder} .

In 530 measurements T_{Blood} was < 36 °C, in 6665 measurements T_{Blood} was 36–38.3 °C and in 470 measurements T_{Blood} was > 38.3 °C.

Bland Altman analysis. Bias between T_{ZHF} and T_{Blood} was -0.12 °C with an upper limit of agreement of 0.51 °C and a lower limit of agreement of -0.76 °C (Fig. 1). Bias between T_{Bladder} and T_{Blood} was 0.05 °C with an upper limit of agreement of 0.64 °C and a lower limit of agreement of -0.54 °C (Fig. 2).

Proportion of differences within ± 0.5 °C and ± 1 °C. The proportion of differences within ± 0.5 °C of T_{Blood} was 90.98% for T_{ZHF} and 95.99% for T_{Bladder} and the proportion of differences within ± 1.0 °C of T_{Blood} was 98.99% for T_{ZHF} and 99.01% for T_{Bladder} .

Sensitivity, specificity, positive and negative predictive values. The calculated sensitivity, specificity, positive and negative predictive values for the detection of hypothermia and fever are shown in Table 1.

	Sensitivity [%]	Specificity [%]	PPV [%]	NPV [%]
Detection of hypothermia				
T_{ZHF}	0.89	0.96	0.62	0.99
$T_{Bladder}$	0.81	0.99	0.84	0.96
Detection of fever				
T_{ZHF}	0.65	0.98	0.74	0.97
$T_{Bladder}$	0.83	0.97	0.67	0.98

Table 1. Sensitivity, specificity, positive and negative predictive values for the detection of hypothermia and fever of both methods. T_{ZHF} temperature measured with a zero-heat flux thermometer, $T_{Bladder}$ bladder temperature, PPV positive predictive value, NPV negative predictive value.




Figure 3. Error grid analysis of the zero-heat flux thermometer (ZHF) against blood temperature (T_{Blood}). Zone A is drawn in white, Zone B in grey and Zone C in dark grey.

Error grid analysis. Error grid analysis showed that 91.6% of all T_{ZHF} measurements were clinically not different from T_{Blood} , or would still lead to the same treatment or diagnostic workup. In 6.2% measurement errors were $>0.5\text{ }^{\circ}\text{C}$, but the result would not lead to a clinical wrong decision. Only 2.2% of the measurements would lead to wrong clinical decisions (Fig. 3). Error grid analysis of $T_{Bladder}$ showed that 96.3% of all measurements were clinically not different from T_{Blood} or would still lead the same treatment or diagnostic workup. In 2.4% measurement errors were $>0.5\text{ }^{\circ}\text{C}$ but this would not result in a clinical wrong decision. Only 1.3% of the measurements would lead to wrong clinical decisions (Fig. 4).

Adverse events. The ZHF-thermometer sensors were well tolerated in all patients and no burn or skin reaction was observed during the study period.

Discussion

In this study with critically ill patients, $T_{Bladder}$ represented T_{Blood} more accurate than T_{ZHF} . In the Bland Altman analyses the bias was smaller and limits of agreement were narrower. The proportion of differences within $\pm 0.5\text{ }^{\circ}\text{C}$ of T_{Blood} were higher, and there were less values in Zone B and C of the error grid analysis. In addition, the ZHF thermometer failed to record core temperatures below $30\text{ }^{\circ}\text{C}$. However, compared to the published results for other non-invasive thermometers like infrared tympanic membrane thermometers, temporal artery thermometers, or axillary thermometers¹² the ZHF-thermometer is more accurate.

Interpretation of our results. The results of the Bland Altman analysis of $T_{Bladder}$ were comparable to the results that were obtained by Nierman²¹ and slightly different from the results of Lefrant et al.²² who observed a bias of $-0.21\text{ }^{\circ}\text{C}$ and more narrow limits of agreement of $\pm 0.20\text{ }^{\circ}\text{C}$. In general, the high level of accuracy of $T_{Bladder}$ is remarkable because oliguria, that is very frequent in ICU patients, reduces the accuracy of $T_{Bladder}$




Figure 4. Error grid analysis of bladder temperature (T_{Bladder}) against blood temperature (T_{Blood}). Zone A is drawn in white, Zone B in grey and Zone C in dark grey.

measurements in operative patients^{23,24}. On the other hand, in critically ill patients, oliguria does not seem to influence the accuracy of bladder temperature very much¹⁰.

The results of the Bland Altman analysis of the ZHF-thermometer were a slightly better than the results that were obtained by Eshraghi et al.¹³ before and after cardiopulmonary bypass, and in the same range as found by Mäkinen et al.¹⁵ during cardiac surgery when the patients were off cardiopulmonary bypass. During surgery with slow temperature changes Boisson et al.¹⁴ could obtain better results with a bias to T_{Blood} of -0.1 °C with limits of agreement of ± 0.4 °C.

The proportion of differences within ± 0.5 °C of T_{Blood} was 84% in the study of Eshraghi¹³ and 94% in the study of Boisson¹⁴. Our results of 91% are also in that range. Other studies that have evaluated the ZHF-thermometer in critically ill patients did not compare it to a gold standard and are therefore of limited value for the comparison with our results^{25,26}.

The question is, if the accuracy of the ZHF-thermometer is still good enough to be used in ICU. Many studies that compare new temperature monitoring devices with a gold standard use a definition that the combined inaccuracy (bias and limits of agreement) should be smaller than 0.5 °C²⁷ to be accurate enough. In our opinion this objective is very high and most of the studies that have investigated new non-invasive thermometers^{13,28,29} did not find an accuracy that met this criterion. Still they came to the conclusion that the new devices agree sufficiently enough for clinical practice^{13,28,29}.

Another possibility is to look at the proportion of differences within the range of ± 0.5 °C of the T_{Blood} . In our study 91% of all measurement values of the ZHF-thermometer were within the range of ± 0.5 °C of T_{Blood} and 99% were within the range of ± 1 °C. That seems to be acceptable.

Another interesting way of interpreting the results is the error grid analysis²⁰. In this analysis 91.6% of the values of the ZHF-thermometer lead to the right clinical decision and only 2.2% of the measured values would lead to wrong clinical decisions. This seems to be sufficient, especially because T_{Core} changes do not require an immediate change in therapy in the next minutes. However, it has been argued, that no single measurement value should be in Zone C as this will lead to wrong clinical decisions²⁰. This seems to be very demanding. If we would accept this, methods like non-invasive blood pressure measurement or pulse oximetry would have to be abandoned immediately.

Limitations of the study

In most studies comparing temperature measurement devices there are many data pairs per subject and the number of data pairs per patient are not equal. This can induce random effects because there are independent influences of the different patients and there are influences of time in each individual patient. This influence is not totally independent. To account for this effect, we have used the Bland and Altman random effects method for repeated measures data adjusted for unequal numbers of measurements per patient¹⁹.

A potential limitation of the methods used is the use of error grid analysis. This method has not been used for the comparison of temperature measurement devices before. Error grid analysis is highly dependent on the zones, which can be by definition arbitrarily defined. In this study the zones were defined by the authors a priori using published and well accepted definitions. Zone A was defined as an area of a ± 0.5 °C error on either side of a perfectly accurate measurement between T_{Blood} and the T_{ZHF} or T_{Bladder} because measurement errors smaller than ± 0.5 °C are considered by most authors as clinically not relevant. In addition, if both measurements indicate

hypothermia $< 36^{\circ}\text{C}$ or fever $> 38.3^{\circ}\text{C}$ the absolute error is considered to be clinically irrelevant because the same treatment or diagnostic workup will be initiated. It can be argued that there is a clinically relevant difference between 35.0°C and 26°C . This would still lead to a data point that is in the Zone A. However, it is extremely difficult to define thresholds for this situation. In addition, we did not observe this.

Other potential limitations of our study are that we have studied a relatively small population with only 50 patients. However, in average every patient was monitored more than 3.3 days, resulting in an average of about 140 measurement points per patient.

Another potential limitation is that we have studied a mixed ICU patient collective. This may also be seen as an advantage because we have measured different patients in different critically ill states and with different influences like renal replacement therapy (RRT) or Extracorporeal Membrane Oxygenation (ECMO). Patients undergoing targeted temperature management after cardiopulmonary resuscitation which might be an interesting and challenging patient cohort in which T_{Core} measurement is of utmost importance are missing in our collective. This may be a limitation to the generalizability of the study results.

In some of our patients the gold standard blood temperature may be distorted by the rapid infusion of unwarmed fluids or extracorporeal devices like RRT or ECMO. It is well known that a rapid infusion of unwarmed or cold fluids can lower blood temperature temporarily. This effect is typically used for the measurement cardiac output with a pulmonary artery catheter. This effect varies depending on the temperature, amount, and rate of the fluid given. Initiation of RRT also temporarily changes blood temperature to a small amount but a stable running RRT does not lead to changes in blood temperature. The same is probably true for ECMO. Infusion of intravenous fluids or RRT are typical measures in ICU and it is not possible to exclude patients that need intravenous fluids. In our patient group 17 patients had RRT and 2 patients had ECMO. This may have contributed to the observed inaccuracy of the ZHF-thermometer and T_{Bladder} . Another potential problem may be that the analogue data transfer from the ZHF-thermometer to the general ICU monitoring may have introduced an additional error.

We did also not observe many measurements for temperature above 39°C . Therefore, it is not possible to make any conclusions about the accuracy the devices in these extremely high temperature range.

The use of vasopressor therapy and especially the use of high dose vasopressor therapy may also influence the accuracy of the ZHF-thermometer. Unfortunately, we did not look at this potential source of inaccuracy. This might be investigated in another study.

Also we did not measure the urine output of our patients, therefore a correlation to accuracy of T_{Bladder} is impossible.

Some studies have used more complex statistical methods²⁹ like population analysis³⁰. However, very often these complex analyses do not add very much new information about the accuracy of the studied devices. We included sensitivity, specificity, positive and negative predictive values for the detection of hypothermia and fever for both methods because this has not been done yet. We also included an error grid analysis because this kind of analysis may be clinically useful although the definition of the three zones in that error grid analysis can be discussed.

Conclusion

In conclusion the results for zero-heat-flux and bladder temperatures were virtually identical within about a tenth of a degree, although T_{ZHF} tended to underestimate T_{Blood} . Therefore, either is suitable for clinical use and can be used if bladder temperature is not available.

Data availability

The datasets used for the analysis in the current study are available from the corresponding author on reasonable request.

Received: 4 April 2020; Accepted: 30 November 2020

Published online: 10 December 2020

References

1. Rohrer, M. J. & Natale, A. M. Effect of hypothermia on the coagulation cascade. *Crit. Care Med.* **20**, 1402–1405 (1992).
2. Hohn, L. *et al.* Benefits from intraoperative skin surface warming in cardiac surgical patients. *Br. J. Anaesth.* **80**, 318–323 (1998).
3. Insler, S. R. *et al.* Association between postoperative hypothermia and adverse outcome after coronary artery bypass surgery. *Ann. Thorac. Surg.* **70**, 175–181 (2000).
4. Kurz, A., Sessler, D. I. & Lenhardt, R. Perioperative normothermia to reduce the incidence of surgical- wound infection and shorten hospitalization. Study of Wound Infection and Temperature Group. *N. Engl. J. Med.* **334**, 1209–1215 (1996).
5. Melling, A. C., Ali, B., Scott, E. M. & Leaper, D. J. Effects of preoperative warming on the incidence of wound infection after clean surgery: a randomised controlled trial. *Lancet* **358**, 876–880 (2001).
6. Scott, A. V. *et al.* Compliance with surgical care improvement Project for body temperature management (SCIP Inf-10) is associated with improved clinical outcomes. *Anesthesiology* **123**, 116–125 (2015).
7. Billeter, A. T. *et al.* Unintentional perioperative hypothermia is associated with severe complications and high mortality in elective operations. *Surgery* **156**, 1245–1252 (2014).
8. Karalapillai, D. *et al.* Inadvertent hypothermia and mortality in postoperative intensive care patients: retrospective audit of 5050 patients. *Anaesthesia* **64**, 968–972 (2009).
9. Niven, D. J., Stelfox, H. T. & Laupland, K. B. Hypothermia in adult ICUs: changing incidence but persistent risk factor for mortality. *J. Intensive Care Med.* **31**, 529–536 (2014).
10. O'Grady, N. P. *et al.* Guidelines for evaluation of new fever in critically ill adult patients: 2008 update from the American College of Critical Care Medicine and the Infectious Diseases Society of America. *Crit. Care Med.* **36**, 1330–1349 (2008).
11. Wartzek, T., Mühlsteff, J. & Imhoff, M. Temperature measurement. *Biomed. Tech. (Berl)* **56**, 241–257 (2011).
12. Niven, D. J. *et al.* Accuracy of peripheral thermometers for estimating temperature: a systematic review and meta-analysis. *Ann. Intern. Med.* **163**, 768–777 (2015).

13. Eshraghi, Y. *et al.* An evaluation of a zero-heat-flux cutaneous thermometer in cardiac surgical patients. *Anesth. Analg.* **119**, 543–549 (2014).
14. Boisson, M. *et al.* Intra-operative cutaneous temperature monitoring with zero-heat-flux technique (3M SpotOn) in comparison with oesophageal and arterial temperature: a prospective observational study. *Eur. J. Anaesthesiol.* **35**, 825–830 (2018).
15. Mäkinen, M. T. *et al.* Novel zero-heat-flux deep body temperature measurement in lower extremity vascular and cardiac surgery. *J. Cardiothorac. Vasc. Anesth.* **30**, 973–978 (2016).
16. Pesonen, E. *et al.* The focus of temperature monitoring with zero-heat-flux technology (3M Bair-Hugger): a clinical study with patients undergoing craniotomy. *J. Clin. Monit. Comput.* **33**, 917–923 (2019).
17. Jansen, T. C., Kompanje, E. J. & Bakker, J. Deferred proxy consent in emergency critical care research: ethically valid and practically feasible. *Crit. Care Med.* **37**, S65–S68 (2009).
18. von Elm, E. *et al.* The Strengthening of Reporting of Observational Studies in Epidemiology (STROBE) statement: guidelines for reporting observational studies. *Lancet* **370**, 1453–1457 (2007).
19. Bland, J. M. & Altman, D. G. Agreement between methods of measurement with multiple observations per individual. *J. Biopharm. Stat.* **17**, 571–582 (2007).
20. Morey, T. E., Gravenstein, N. & Rice, M. J. Let's think clinically instead of mathematically about device accuracy. *Anesth. Analg.* **113**, 89–91 (2011).
21. Nierman, D. M. Core temperature measurement in the intensive care unit. *Crit. Care Med.* **19**, 818–823 (1991).
22. Lefrant, J.-Y. *et al.* Temperature measurement in intensive care patients: comparison of urinary bladder, oesophageal, rectal, axillary, and inguinal methods versus pulmonary artery core method. *Intensive Care Med.* **29**, 414–418 (2003).
23. Sato, H. *et al.* Urinary bladder and oesophageal temperatures correlate better in patients with high rather than low urinary flow rates during non-cardiac surgery. *Eur. J. Anaesthesiol.* **25**, 805–809 (2008).
24. Horrow, J. C. & Rosenberg, H. Does urinary catheter temperature reflect core temperature during cardiac surgery?. *Anesthesiology* **69**, 986–979 (1988).
25. Schell-Chaple, H. M., Liu, K. D., Matthay, M. A. & Puntillo, K. A. Rectal and bladder temperatures vs forehead core temperatures measured with SpotOn monitoring system. *Am. J. Crit. Care.* **27**, 43–50 (2018).
26. Dahyot-Fizelier, C. *et al.* Accuracy of zero-heat-flux cutaneous temperature in intensive care adults. *Crit. Care Med.* **45**, e715–e717 (2017).
27. Sessler, D. I. Temperature monitoring and perioperative thermoregulation. *Anesthesiology* **109**, 318–338 (2008).
28. Kimberger, O. *et al.* Accuracy and precision of a novel non-invasive core thermometer. *Br. J. Anaesth.* **103**, 226–231 (2009).
29. Kimberger, O. *et al.* The accuracy of a disposable noninvasive core thermometer. *Can. J. Anaesth.* **60**, 1190–1196 (2013).
30. Soehle, M., Dehne, H., Hoeft, A. & Zenker, S. Accuracy of the non-invasive Tcore temperature monitoring system to measure body core temperature in abdominal surgery. *J. Clin. Monit. Comput.* **34**, 1361–1367 (2020).

Acknowledgements

We would like to thank the nursing staff for supporting the study.

Author contributions

A.B., A.F., T.P., and I.F.B. designed the study. A.B., A.F., D.H. and K.M. managed data and its quality. A.B., A.F., and I.F.B. performed the statistical analysis. All authors participated in the data interpretation. A.B. drafted the manuscript. A.F., D.H., T.P., I.F.B. and K.M. contributed substantially to its revision. All authors read the manuscript carefully and approved the final version.

Funding

Open Access funding enabled and organized by Projekt DEAL. The study was conducted with departmental funding only. No equipment was loaned or supplied by a company.

Competing interests

Prof. Dr. A. Bräuer is a member of the advisory board of 3 M Europe and has received payments from 3 M Germany, 3 M Europe, 3 M Asia Pacific Pte Ltd. for consultancy work. PD Dr. T. Perl T has received consulting honorary from The 37Company the Netherlands and Barkey, Germany. The other authors declare no competing interests.

Additional information

Correspondence and requests for materials should be addressed to A.B.

Reprints and permissions information is available at www.nature.com/reprints.

Publisher's note Springer Nature remains neutral with regard to jurisdictional claims in published maps and institutional affiliations.



Open Access This article is licensed under a Creative Commons Attribution 4.0 International License, which permits use, sharing, adaptation, distribution and reproduction in any medium or format, as long as you give appropriate credit to the original author(s) and the source, provide a link to the Creative Commons licence, and indicate if changes were made. The images or other third party material in this article are included in the article's Creative Commons licence, unless indicated otherwise in a credit line to the material. If material is not included in the article's Creative Commons licence and your intended use is not permitted by statutory regulation or exceeds the permitted use, you will need to obtain permission directly from the copyright holder. To view a copy of this licence, visit <http://creativecommons.org/licenses/by/4.0/>.

© The Author(s) 2020