

RESEARCH

Open Access



Factors influencing the accuracy of multimodal image fusion for oral and maxillofacial tumors: a retrospective study

Lei-Hao Hu^{1†}, Wen-Bo Zhang^{1†}, Yao Yu¹, Zhi-Peng Sun², Guang-Yan Yu¹ and Xin Peng^{1*}

Abstract

Background: Ensuring high accuracy in multimodal image fusion for oral and maxillofacial tumors is crucial before further application. The aim of this study was to explore the factors influencing the accuracy of multimodal image fusion for oral and maxillofacial tumors.

Methods: Pairs of single-modality images were obtained from oral and maxillofacial tumor patients, and were fused using a proprietary navigation system by using three algorithms (automatic fusion, manual fusion, and registration point-based fusion). Fusion accuracy was evaluated including two aspects—overall fusion accuracy and tumor volume fusion accuracy—and were indicated by mean deviation and fusion index, respectively. Image modality, fusion algorithm, and other characteristics of multimodal images that may have potential influence on fusion accuracy were recorded. Univariate and multivariate analysis were used to identify relevant affecting factors.

Results: Ninety-three multimodal images were generated by fusing 31 pairs of single-modality images. The interaction effect of image modality and fusion algorithm ($P=0.02$, $P=0.003$) and thinner slice thickness ($P=0.006$) were shown to significantly influence the overall fusion accuracy. The tumor volume ($P<0.001$), tumor location ($P=0.007$), and image modality ($P=0.01$) were significant influencing factors for tumor volume fusion accuracy.

Conclusions: To ensure high overall fusion accuracy, manual fusion was not preferred in CT/MRI image fusion, and neither was automatic fusion in image fusion containing PET modality. Using image sets with thinner slice thickness could increase overall fusion accuracy. CT/MRI fusion yielded higher tumor volume fusion accuracy than fusion containing PET modality. The tumor volume fusion accuracy should be taken into consideration during image fusion when the tumor volume is small and the tumor is located in the mandible.

Keywords: Oral and maxillofacial tumors, Multimodal image fusion, Accuracy, Influencing factors

[†]Lei-Hao Hu and Wen-Bo Zhang contributed equally to the work and share first authorship

*Correspondence: pxpengxin@263.net

¹ Department of Oral and Maxillofacial Surgery, Peking University School and Hospital of Stomatology & National Center of Stomatology & National Clinical Research Center for Oral Diseases & National Engineering Research Center of Oral Biomaterials and Digital Medical Devices, No.22, Zhongguancun South Avenue, Haidian District, Beijing, People's Republic of China

Full list of author information is available at the end of the article

Background

Multimodal image fusion, which provides different modalities of images integrated into a common reference frame under specific algorithms and displayed in one pair of multimodal images, has been frequently used in the diagnosis, virtual surgical planning, radiotherapeutic planning, and follow-up for oral and maxillofacial tumors [1–8]. Single-modality image sets like computed tomography (CT), magnetic resonance imaging (MRI), and positron-emission tomography (PET) can be registered and



fused into one multimodal image set, which could reveal considerably more information regarding tumor infiltration and the spatial relativity between tumor and surrounding tissues than single-modality images.

The quality assurance of multimodal image fusion is the foundation of its in-depth application, and the core issue of quality assurance is ensuring high fusion accuracy [9]. Despite the wide application of multimodal image fusion, only a few of studies have focused on the accuracy of multimodal image fusion for oral and maxillofacial tumors. Most studies have demonstrated a fusion accuracy of ≤ 2 mm [9–13]. A previous study proposed a revised method to evaluate fusion accuracy that included the overall fusion accuracy—represented by mean deviation (MD) of six pairs of landmark points—and the tumor volume fusion accuracy—represented by Fusion Index (FI) [14]. The MD value ranged from 1.926 to 2.788 mm for different fusion algorithms, which showed a similar result of overall fusion accuracy with former studies. The FI value ranged from 0.520 to 0.594, and it was a newly proposed indicator of fusion accuracy that revealed to what degree the volume of the tumor on different modalities of image overlapped.

To achieve a relatively high accuracy of diagnosis or treatment planning by using multimodal image fusion, the fusion accuracy needs to be assured. Some researchers reported that the accuracy of multimodal image fusion was influenced by certain factors such as the parameters of single-modality image (i.e., pixel pitch and slice thickness) and the patients' status when being scanned (i.e., the patient position and organ movement) [15, 16]. A previous study reported that multimodal image fusion that contained PET images seemed less accurate than CT/MRI fusion, which implied that the modality may be an influencing factor for fusion accuracy [14]. To our knowledge, no study has yet evaluated the potential influencing factors of the accuracy of multimodal image fusion for oral and maxillofacial tumors.

Ensuring high accuracy in multimodal image fusion for oral and maxillofacial tumors is crucial before further application. Therefore, this study explored the factors that influence the accuracy of multimodal image fusion for oral and maxillofacial tumors.

Methods

Patients and single-modality images

This study is a retrospective study and enrolled patients with a diagnosis of an oral and maxillofacial tumor who were referred to our department from January 2019 to January 2020. The inclusion criteria were as follows: (1) patients in whom the tumor was located in a deep oral area (e.g., gingiva of the posterior teeth, soft palate)

or deep maxillofacial area (e.g., maxillary sinus, skull base, infratemporal fossa), and infiltrated at least two anatomical regions; (2) patients who had undergone at least two modalities of radiologic examination pre-operatively and for whom complete Digital Imaging and Communications in Medicine (DICOM) files of at least two imaging modalities among regular CT (or contrast-enhanced CT, ceCT), MRI (T2 weighted or contrast-enhanced fat-suppressed T1 weighted), and PET-CT including maxillofacial area were available. The exclusion criteria were: (1) patients for whom the time interval between different radiological scans were over 20 days, as this might have led to tissue deformation caused by tumor growth; (2) the parameters of the patients' image scans were ambiguous or could not be acquired from the DICOM files [14]. There was an overlapping in the patients enrolled in this study with the previously published works of the same research team, and this study expanded the sample sizes on that basis [14].

The study was approved by the Biomedical Institutional Review Board of Peking University School of Stomatology (approval number: PKUSSIRB-202054021).

Multimodal image fusion

The DICOM files of single-modality image sets of the same patient were imported into iPlan CMF 3.0 (BrainLAB, Feldkirchen, Germany) (Fig. 1a). By using “image fusion” module, two single-modality image sets were fused into one multimodal image set. The fusion modalities included two types: (1) CT/MRI image fusion (CT or ceCT image sets fused with MRI image sets); (2) PET-containing image fusion (PET-CT image sets fused with ceCT image sets, or PET-CT image sets fused with MRI image sets) (Fig. 1b). Three fusion algorithms were applied for every two single-modality image sets: (1) automatic fusion, finished automatically by the iPlan CMF software under the principle of maximization of mutual information in the region of interest; (2) manual fusion, finished manually by operating staff through translating or rotating one image set to align the other image set as much as possible; (3) registration point-based fusion: finished by iPlan CMF software through matching the corresponding registration points on two image sets that were marked by the operating staff before image fusion, and the registration points were all anatomical landmarks located at maxillofacial regions.

The process of image fusion was completed after two oral and maxillofacial surgeons (W.-B. Z., Y. Y.) with 8-year-experience in using the iPlan CMF software reached consensus.




Fig. 1 The schematic diagram of multimodal image fusion. **a** Three types of single-modality image sets were enrolled in this study. **b** The results of multimodal image fusion, which could be classified into CT/MRI image fusion and PET-containing image fusion according to the image modalities

Evaluating the fusion accuracy

The accuracy of multimodal image fusion was evaluated for every fusion project, including overall fusion accuracy and tumor volume fusion accuracy (Fig. 2) [14].

To evaluate overall fusion accuracy, six pairs of anatomical landmarks were marked correspondingly on two image sets, three-dimensionally representing the superior, inferior, anterior, posterior, left, and right boundary of the overall image sets: (A) the interior point of nasion; (B) the tangency point of the upper central incisors; (C) the former point of the anterior nasal spine; (D) the former point of the atlas; (E) the tangency point of the left mandibular notch; (F) the tangency point of the right mandibular notch (Fig. 2a). The operating staff recorded the three-dimensional coordinates of six pairs of landmarks on two single-modality image sets under the same coordinate system. The coordinates of points on one image set were (x_{i1}, y_{i1}, z_{i1}) ($i = 1, 2, \dots, 6$), while those on the other image set were (x_{i2}, y_{i2}, z_{i2}) ($i = 1, 2, \dots, 6$) (Fig. 2b). The mean values of the coordinate differences of six pairs of landmarks along the x-, y-, and z-axes were Δx , Δy , and Δz , which were calculated as shown: $\Delta x = \left| \frac{\sum_{i=1}^6 (x_{i1} - x_{i2})}{6} \right|$,

$$\Delta y = \left| \frac{\sum_{i=1}^6 (y_{i1} - y_{i2})}{6} \right|, \quad \Delta z = \left| \frac{\sum_{i=1}^6 (z_{i1} - z_{i2})}{6} \right|.$$

The overall fusion accuracy was represented by MD, which was calculated as shown: $MD = \sqrt{\Delta x^2 + \Delta y^2 + \Delta z^2}$. The MD

value revealed the overall deviation of the same landmark points on two image sets. The lower the value of MD, the less the deviation between two image sets, and the better the overall fusion accuracy.

To evaluate tumor volume fusion accuracy, the tumor was delineated separately on two image sets, then the tumor volume on each image set (V_F and V_M , presented in red color and in blue color) and intersected part of the tumor (V_{F+M} , presented in green color) were generated automatically (Fig. 2c). The tumor volume fusion accuracy was indicated by FI, which was calculated as shown: $FI = \frac{V_{F+M}}{V_F} \times \frac{V_{F+M}}{V_M}$. The higher the value of FI, the more the intersected part of tumor volume, and the better the tumor volume fusion accuracy.

The overall fusion accuracy and the tumor volume fusion accuracy were both evaluated twice by a well-experienced oral and maxillofacial surgeon (L.-H. H.) with 5-year-experience in using the iPlan CMF software and not participating in the image fusion process. The final result of fusion accuracy was the mean value of the two results. The evaluation process was finished under the guidance of a board-certified radiologist (Z.-P. S.) with 18-year-experience in radiological diagnosing of oral and maxillofacial tumors.

Collection of characteristics and classification of variables

Patient and imaging characteristics were classified as follows: (1) Categorical variables including nature of




Fig. 2 The schematic diagram of evaluating the fusion accuracy. **a** The location of the six pairs of anatomical landmarks. **b** The process of evaluating overall fusion accuracy which was represented as MD. **c** The process of evaluating the tumor volume fusion accuracy which was represented as FI

the tumor, tumor location, dental artifact, change of patient's position among different image scans, fusion modality, fusion algorithm and (2) Numerical variables including gross tumor volume, slice thickness, and pixel pitch of single-modality images. The value of gross tumor volume was acquired by averaging the tumor volume on different image sets that were generated automatically by evaluating tumor volume fusion accuracy.

Statistical analysis

All measured data were analyzed using SPSS Statistics v24.0 (IBM Corp., Armonk, NY). MD and FI were set as the dependent variables. Univariate analysis was first carried out, including analysis of variance (ANOVA) for categorical variables and correlation analysis for numerical variables. Factors with $P < 0.10$ in the univariate analysis were included in the multivariate linear regression models to identify significant influencing factors. $P < 0.05$ in multivariate analysis was considered to indicate statistical significance.

Results

Overview of multimodal image fusion

Ninety-three multimodal image sets were generated by fusing 31 pairs of single-modality image sets. The basic characteristics of these 93 multimodal image sets are shown in Table 1.

Univariate analysis

The results of univariate analysis are shown in Tables 2 and 3. Change of patient's position among different image scans ($P = 0.05$) and thinner slice thickness ($P = 0.01$) were shown to potentially influence the overall fusion accuracy. Fusion modality ($P < 0.001$), nature and location of the tumor (both $P < 0.001$), and higher pixel pitch ($P = 0.09$) potentially influence the tumor volume fusion accuracy.

Per one-way ANOVA, neither the fusion modality nor the fusion algorithm was potential influencing factors of overall fusion accuracy, as their P -values were > 0.10 . Nevertheless, the variation trend of the overall fusion accuracy of different fusion algorithms was opposite

Table 1 Basic characteristics of 93 multimodal images

Categories	Amount/Mean $\pm$ SD
Modalities	
CT/MRI	48
PET-containing	45
Fusion algorithm	
Automatic fusion	31
Manual fusion	31
Registration point-based fusion	31
Nature of tumor	
Benign	30
Malignant	63
Location of tumor	
Maxilla	57
Mandible	36
Dental artifacts	
Absent	78
Present	15
Change of patients' position	
Absent	60
Present	33
Gross tumor volume (cm ³)	31.67 $\pm$ 31.62
Interval days between different scans	7.76 $\pm$ 6.36
Thinner slice thickness (mm)	1.57 $\pm$ 0.70
Thicker slice thickness (mm)	3.03 $\pm$ 1.18
Lower pixel pitch (mm)	0.44 $\pm$ 0.13
Higher pixel pitch (mm)	1.92 $\pm$ 1.62

SD Standard deviation

among different fusion modalities: for the CT/MRI modality, the overall fusion accuracy of automatic fusion was the best among three fusion algorithms, but it could be the worst when it comes to fusion modality containing PET. Such trend was not observed in tumor volume fusion accuracy (Fig. 3). This phenomenon implied that there was an interaction effect among fusion modalities and fusion algorithms, which could influence the overall fusion accuracy.

Therefore, two-way ANOVA was additionally performed, which verified the previous assumption: the interaction effect of modality and fusion algorithm was a potential influencing factor of overall fusion accuracy ($P < 0.001$) and would not influence the tumor volume fusion accuracy ($P = 0.98$) (Fig. 3).

Multivariate analysis

Overall fusion accuracy

The model for multivariate linear regression analysis in which MD was set as the dependent variable was constructed using the variables that were significant in the univariate analysis (interaction effect of modality and

Table 2 Univariate analysis of potential influencing factors (categorical variables) of fusion accuracy

Characteristics	MD (mm)	FI
Modalities		
CT/MRI	2.06 $\pm$ 1.27	0.61 $\pm$ 0.14
PET-Containing	2.35 $\pm$ 2.17	0.41 $\pm$ 0.16
<i>P</i> -value	0.45	< 0.001*
Fusion algorithms		
Automatic fusion	2.63 $\pm$ 2.37	0.54 $\pm$ 0.17
Manual fusion	2.02 $\pm$ 1.33	0.48 $\pm$ 0.18
Registration point-based fusion	1.94 $\pm$ 1.35	0.50 $\pm$ 0.18
<i>P</i> -value	0.24	0.43
Location of tumor		
Maxilla	2.25 $\pm$ 1.33	0.63 $\pm$ 0.15
Mandible	2.17 $\pm$ 1.94	0.45 $\pm$ 0.16
<i>P</i> -value	0.81	< 0.001*
Nature of tumor		
Benign	2.05 $\pm$ 1.29	0.56 $\pm$ 0.16
Malignant	2.43 $\pm$ 2.32	0.42 $\pm$ 0.17
<i>P</i> -value	0.38	< 0.001*
Dental artifacts		
Absent	1.97 $\pm$ 1.55	0.55 $\pm$ 0.13
Present	2.24 $\pm$ 1.80	0.50 $\pm$ 0.19
<i>P</i> -value	0.59	0.37
Change of patients' position		
Absent	1.89 $\pm$ 1.27	0.51 $\pm$ 0.14
Present	2.76 $\pm$ 2.32	0.50 $\pm$ 0.23
<i>P</i> -value	0.05*	0.79

* $P < 0.10$

fusion algorithm, change of patient's position among different image scans, and thinner slice thickness) (Table 4). In the model, the P -value of the interaction effect of modality and fusion algorithm was < 0.05 , which meant that the interaction effect among fusion modalities and fusion algorithms had significant influence on the overall fusion accuracy, manifested as:


- (1) Setting the overall fusion accuracy of automatic fusion in CT/MRI modality as the benchmark, both CT/MRI manual fusion ($P = 0.02$) and automatic fusion in the PET-containing modality ($P = 0.003$) could significantly decrease the overall fusion accuracy.
- (2) The overall fusion accuracy could be lower as the thinner slice thickness increased ($P = 0.006$).

Tumor volume fusion accuracy

We set FI as the dependent variable of the model for multivariate linear regression analysis of tumor volume accuracy, while variables that were significant in

Table 3 Univariate analysis of potential influencing factors (numerical variables) of fusion accuracy

	Characteristics	Gross tumor volume	Thinner slice thickness	Thicker slice thickness	Lower pixel pitch	Higher pixel pitch
MD	Pearson's correlation coefficient	−0.10	0.27	0.11	−0.12	−0.10
	P-value	0.33	0.01*	0.32	0.26	0.34
FI	Pearson's correlation coefficient	0.36	−0.01	−0.07	0.18	−0.15
	P-value	< 0.001*	0.90	0.50	0.09*	0.15

* $P < 0.10$ **Table 4** Multivariate analysis of potential influencing factors of overall fusion accuracy

Factors	β (95% CI)	Standardized beta coefficient	P-value
The interaction effect of modality and fusion algorithm			
CT/MRI $\times$ AF			
CT/MRI $\times$ MF	1.25 (0.33–2.17)	0.34	0.02*
CT/MRI $\times$ RPBF	0.43 (−0.49 to 1.34)	0.12	0.40
PET-containing $\times$ AF	1.78 (0.81–2.75)	0.44	0.003*
PET-containing $\times$ MF	−0.30 (−1.27 to 0.67)	−0.07	0.10
PET-containing $\times$ RPBF	0.29 (−0.68 to 1.26)	0.07	0.70
Change of patient's position	0.49 (−0.18 to 1.15)	0.13	0.15
Thinner slice thickness	0.69 (0.21–1.17)	0.29	0.006*

CI Confidence interval, AF Automatic fusion, MF Manual fusion, RPBF Registration point-based fusion;

* $P < 0.05$

the univariate analysis (modality, nature and location of tumor, gross tumor volume, and lower pixel pitch of single-modality images) were included as the independent variables (Table 5). In the model, the P -values of the modality, tumor location, and gross tumor volume were all < 0.05 , showing that these factors could significantly influence the tumor volume fusion accuracy, manifested as:

- (1) The tumor volume fusion accuracy of the CT/MRI modality was better than modality containing PET ($P = 0.01$).
- (2) The tumor volume fusion accuracy was better in tumors located in the maxilla than in the mandible ($P = 0.007$).
- (3) The bigger the gross tumor volume, the higher the tumor volume fusion accuracy ($P < 0.001$).

Table 5 Multivariate analysis of potential influencing factors of tumor volume fusion accuracy

Factors	β (95% CI)	Standardized Beta coefficient	P-value
Modality			
CT/MRI			
PET-containing	-0.12 (-0.20 to -0.03)	-0.33	0.01*
Gross tumor volume	0.002 (0.002–0.003)	0.43	<0.001*
Location of tumor			
Maxilla			
Mandible	-0.10 (-0.17 to -0.03)	-0.27	0.007*
Nature of tumor			
Benign			
Malignant	-0.05 (-0.13 to 0.03)	-0.13	0.21
Lower pixel pitch	-0.14 (-0.35 to 0.07)	-0.12	0.20

CI Confidence interval

* $P < 0.05$

Discussion

This study explored the influencing factors of accuracy in multimodal image fusion for oral and maxillofacial tumors based on 93 multimodal images. The interaction effect of modality and fusion algorithm significantly influenced the overall fusion accuracy ($P < 0.001$). Manual fusion was not recommended in case of CT/MRI image fusion to obtain an accurately fused multimodal image, and automatic fusion was not recommended for PET-containing fusion.

When manual fusion was applied on CT/MRI image fusion, the operator needed to manually translate or rotate one pair of single-modality image set to align with another pair of single-modality image set in the region of interest. Automatic fusion and registration point-based fusion could decrease the errors resulting from determining the destination of translation or rotation manually by the operator in manual fusion, and therefore enhanced the overall fusion accuracy of multimodal image fusion.

According to the user's manual of BrainLAB iPlan CME, the principle of automatic fusion was maximization of mutual information. The software attained the fusion project with the largest mutual information as the final result of the automatic fusion [17, 18]. PET is a modality that reflects the metabolism of tissues and organs through the uptake of tracers in different parts of the body and displays tumor size, volume, and metabolic activity from the perspective of tumor metabolism. Unlike anatomical modalities such as CT or MRI, PET usually owns fewer gray levels, which makes it difficult to differentiate the anatomical structure outside the area of

tumor volume. This can interfere with the calculation of mutual information of software platform, and decrease the mutual information, thereby resulting in a relatively low overall fusion accuracy [18, 19].

The tumor volume is a significant influencing factor of tumor volume fusion accuracy. When the operator carries out multimodal image fusion for small tumors, the tumor volume fusion accuracy should be paid more attention to. If the tumor volume fusion accuracy is not satisfied, fine adjustment of the location of image sets could be made based on aligning the tumor volume to different single-modality image sets.

The results of this study showed that reducing slice thickness could significantly improve the overall fusion accuracy of multimodal image fusion. Besides, pixel pitch was not a significant factor influencing the accuracy of multimodal image fusion. Some studies conducted preliminary investigations on the effect of spatial resolution on the accuracy of multimodal image fusion, and their conclusions were not the same. Ng et al. [20] used phantoms to explore the accuracy of transrectal ultrasound and cone-beam CT (CBCT) multimodal image fusion, suggesting that the slice thickness was not a significant factor influencing the accuracy. Kanakavelu et al. [21] used phantom and patient image data to verify the accuracy of the automatic fusion of kilovolt CT and megavolt CBCT, revealing that higher accuracy could be achieved when the slice thickness was 1 mm. Yang et al. [22] believed that in order to improve the accuracy of CT/MRI multimodal image fusion, CT and MRI image sets should be obtained within thin and consistent slice thickness. Based on the results of this study and previous studies, we could conclude that in multimodal image fusion containing certain kinds of image modality, image sets with thin slice thickness might be conducive to improving the overall fusion accuracy of multimodal image fusion compared to those with thick slice thickness. Nevertheless, it's not clear whether decreasing the slice thickness could improve the accuracy of multimodal image fusion for all kinds of image modality, and it needed to be explored in the future.

Besides, the tumor volume fusion accuracy of the tumor located in the mandible was lower than that of the tumor located in the maxilla. The mandible is a movable structure. Because of malocclusion or edentulous jaws, it could not be guaranteed in some patients whether the upper and lower teeth were in the median position when undergoing different imaging scans, which resulted in the change of spatial position of tumors located in the lower jaw among different modalities of image sets and therefore reduced the tumor volume fusion accuracy [23].

Nevertheless, the sample size of this study was relatively small, and the influencing factors were obtained

based on statistical methods, while the question regarding the mechanism involved in the influence of these factors on fusion accuracy remains largely unanswered, and further studies are needed.

Conclusions

This study explored the influencing factors of the accuracy of multimodal image fusion for oral and maxillofacial tumors. To assure high overall fusion accuracy, manual fusion was not preferred in CT/MRI image fusion, and neither was automatic fusion in PET-containing image fusion. Using image sets with thinner slice thickness could increase the overall fusion accuracy. CT/MRI fusion yielded higher tumor volume fusion accuracy than PET-containing fusion. The tumor volume fusion accuracy should be taken into consideration during image fusion when the tumor volume is small and the tumor is located in the mandible.

Abbreviations

CT: Computed tomography; MRI: Magnetic resonance imaging; PET: Positron emission tomography; MD: Mean deviation; FI: Fusion index; DICOM: Digital Imaging and Communications in Medicine; ceCT: Contrast enhanced computed tomography; ANOVA: Analysis of variance; CBCT: Cone-beam computed tomography; AF: Automatic fusion; MF: Manual fusion; RPBF: Registration point-based fusion.

Acknowledgements

We appreciate the professional editor at Elixigen for revising and modifying the English language of this manuscript.

Author contributions

L.-H. H. and W.-B. Z. collected and analyzed the data, and draft the article. Y. Y. and Z.-P. S. collected the data. L.-H. H. statistically analyzed the data. W.-B. Z. and X. P. conceived the ideas and study design, and critically revised and approved the article. G.-Y. Y. and X. P. revised and approved the article. All authors gave final approval and agree to be accountable for all aspects of the work. L.-H. H. and W.-B. Z. contributed equally to the work and share first authorship. All authors read and approved the final manuscript.

Funding

This work was supported by Program of the new clinical techniques of Peking University School and Hospital of Stomatology (Grant Number: PKUSSNCT-20A05).

Availability of data and materials

All data and materials as well as software application or custom code are available from the corresponding author on reasonable request.

Declarations

Ethics approval and consent to participate

All procedures performed in studies involving human participants were in accordance with the ethical standards of the institutional and/or national research committee and with the 1964 Helsinki Declaration and its later amendments or comparable ethical standards. The study was approved by the Biomedical Institutional Review Board of Peking University School of Stomatology (approval number: PKUSSIRB-202054021). Written informed consent was not required for this study because all of the included patients in the present investigation were collected retrospectively. All of the patients enrolled in this study have received treatment in our department, and they have signed informed consent form before treatment, where they consented that their image data could be acquired for research use in our department. Based on

these reasons, the application for free informed consent has been approved by the Biomedical Institutional Review Board of Peking University School of Stomatology. Exemption of informed consent will not affect the rights and health of included patients.

Consent for publication

Not applicable.

Competing interests

The authors declare that they have no competing interests.

Author details

¹Department of Oral and Maxillofacial Surgery, Peking University School and Hospital of Stomatology & National Center of Stomatology & National Clinical Research Center for Oral Diseases & National Engineering Research Center of Oral Biomaterials and Digital Medical Devices, No.22, Zhongguancun South Avenue, Haidian District, Beijing, People's Republic of China. ²Department of Oral and Maxillofacial Radiology, Peking University School and Hospital of Stomatology & National Center of Stomatology & National Clinical Research Center for Oral Diseases & National Engineering Research Center of Oral Biomaterials and Digital Medical Devices, Haidian District, Beijing, People's Republic of China.

Received: 29 June 2022 Accepted: 19 December 2022

Published online: 30 December 2022

References

- Leong JL, Batra PS, Citardi MJ. CT-MR image fusion for the management of skull base lesions. *Otolaryngol Head Neck Surg.* 2006;134(5):868–76.
- Feichtinger M, Aigner RM, Karcher H. F-18 positron emission tomography and computed tomography image-fusion for image-guided detection of local recurrence in patients with head and neck cancer using a 3-dimensional navigation system: a preliminary report. *J Oral Maxillofac Surg.* 2008;66(1):193–200.
- Feichtinger M, Pau M, Zemmann W, Aigner RM, Karcher H. Intraoperative control of resection margins in advanced head and neck cancer using a 3D-navigation system based on PET/CT image fusion. *J Craniomaxillofac Surg.* 2010;38(8):589–94.
- Kraeima J, Dorgelo B, Gulbitt HA, Steenbakkens R, Schepman KP, Roodenburg JLN, Spijkervet FKL, Schepers RH, Witjes MJH. Multi-modality 3D mandibular resection planning in head and neck cancer using CT and MRI data fusion: a clinical series. *Oral Oncol.* 2018;81:22–8.
- Zrnc TA, Wallner J, Zemmann W, Pau M, Gstettner C, Brdic L, Assaf AT, Hassanzadeh H, Feichtinger M, Schwenzer-Zimmerer K. Assessment of tumor margins in head and neck cancer using a 3D-navigation system based on PET/CT image-fusion: a pilot study. *J Craniomaxillofac Surg.* 2018;46(4):617–23.
- Loeffelbein DJ, Souvatzoglou M, Wankel V, Dinges J, Ritschl LM, Mücke T, Pickhard A, Eiber M, Schwaiger M, Beer AJ. Diagnostic value of retrospective PET–MRI fusion in head-and-neck cancer. *BMC Cancer.* 2014;14(1):846.
- Queiroz MA, Huellner MW. PET/MR in cancers of the head and neck. *Semin Nucl Med.* 2015;45(3):248–65.
- Liu Y. FDG PET/CT for metastatic squamous cell carcinoma of unknown primary of the head and neck. *Oral Oncol.* 2019;92:46–51.
- Mutic S, Dempsey JF, Bosch WR, Low DA, Drzymala RE, Chao KSC, Goddard SM, Cutler PD, Purdy JA. Multimodality image registration quality assurance for conformal three-dimensional treatment planning. *Int J Radiat Oncol Biol Phys.* 2001;51(1):255–60.
- Daisne J-F, Sibomana M, Bol A, Cosnard G, Lonneux M, Grégoire V. Evaluation of a multimodality image (CT, MRI and PET) coregistration procedure on phantom and head and neck cancer patients: accuracy, reproducibility and consistency. *Radiother Oncol.* 2003;69(3):237–45.
- Ulin K, Urie MM, Cherlow JM. Results of a multi-institutional benchmark test for cranial CT/MR image registration. *Int J Radiat Oncol Biol Phys.* 2010;77(5):1584–9.
- Al-Saleh MA, Punithakumar K, Jaremko JL, Alsufyani NA, Boulanger P, Major PW. Accuracy of magnetic resonance imaging-cone beam

- computed tomography rigid registration of the head: an in-vitro study. *Oral Surg Oral Med Oral Pathol Oral Radiol.* 2016;121(3):316–21.
13. Wang XS, Li LG, Hu CS, Qiu JJ, Xu ZY, Feng Y. A comparative study of three CT and MRI registration algorithms in nasopharyngeal carcinoma. *J Appl Clin Med Phys.* 2009;10(2):3–10.
 14. Hu LH, Zhang WB, Yu Y, Peng X. Accuracy of multimodal image fusion for oral and maxillofacial tumors: a revised evaluation method and its application. *J Craniomaxillofac Surg.* 2020;48(8):741–50.
 15. Lee MW, Park HJ, Kang TW, Ryu J, Bang WC, Lee B, Lee ES, Choi BI. Image fusion of real-time ultrasonography with computed tomography: factors affecting the registration error and motion of focal hepatic lesions. *Ultrasound Med Biol.* 2017;43(9):2024–32.
 16. Ketcha MD, De Silva T, Han R, Uneri A, Goerres J, Jacobson MW, Vogt S, Kleinszig G, Siewerdsen JH. Effects of image quality on the fundamental limits of image registration accuracy. *IEEE Trans Med Imaging.* 2017;36(10):1997–2009.
 17. Maes F, Collignon A, Vandermeulen D, Marchal G, Suetens P. Multimodality image registration by maximization of mutual information. *IEEE Trans Med Imaging.* 1997;16(2):187–98.
 18. Marinelli M, Positano V, Tucci F, Neglia D, Landini L. Automatic PET-CT image registration method based on mutual information and genetic algorithms. *Sci World J.* 2012;2012:567067.
 19. Klabbers BM, de Munck JC, Slotman BJ, Langendijk HA, de Bree R, Hoekstra OS, Boellaard R, Lammertsma AA. Matching PET and CT scans of the head and neck area: development of method and validation. *Med Phys.* 2002;29(10):2230–8.
 20. Ng A, Beiki-Ardakan A, Tong S, Moseley D, Siewerdsen J, Jaffray D, Yeung IW. A dual modality phantom for cone beam CT and ultrasound image fusion in prostate implant. *Med Phys.* 2008;35(5):2062–71.
 21. Kanakavelu N, Samuel EJ. Accuracy in automatic image registration between MV cone beam computed tomography and planning kV computed tomography in image guided radiotherapy. *Rep Pract Oncol Radiother.* 2016;21(5):487–94.
 22. Yang R, Li QX, Mao C, Peng X, Wang Y, Guo YX, Guo CB. Multimodal image fusion technology for diagnosis and treatment of the skull base-infratemporal tumors. *J Peking Univ (Health Sci).* 2019;51(1):53–8.
 23. Hanvey S, Sadozye AH, McJury M, Glegg M, Foster J. The influence of MRI scan position on image registration accuracy, target delineation and calculated dose in prostatic radiotherapy. *Br J Radiol.* 2012;85(1020):e1256–1262.

Publisher's Note

Springer Nature remains neutral with regard to jurisdictional claims in published maps and institutional affiliations.

Ready to submit your research? Choose BMC and benefit from:

- fast, convenient online submission
- thorough peer review by experienced researchers in your field
- rapid publication on acceptance
- support for research data, including large and complex data types
- gold Open Access which fosters wider collaboration and increased citations
- maximum visibility for your research: over 100M website views per year

At BMC, research is always in progress.

Learn more biomedcentral.com/submissions

